

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
 Data File : VX043932.D
 Acq On : 21 Nov 2024 13:57
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX112124

Quant Time: Nov 22 03:54:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.633	0.640	-1.1	89	0.00
3 P	Chloromethane	0.667	0.655	1.8	89	0.00
4 C	Vinyl Chloride	0.709	0.711	-0.3#	88	0.00
5 T	Bromomethane	0.436	0.410	6.0	87	0.00
6 T	Chloroethane	0.432	0.354	18.1	73	0.00
7 T	Trichlorofluoromethane	1.268	1.354	-6.8	94	0.00
8 T	Diethyl Ether	0.462	0.437	5.4	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.605	0.614	-1.5	88	0.00
10 T	Methyl Iodide	0.755	0.734	2.8	86	0.00
11 T	Tert butyl alcohol	0.138	0.131	5.1	86	0.00
12 CM	1,1-Dichloroethene	0.576	0.555	3.6#	88	0.00
13 T	Acrolein	0.145	0.137	5.5	89	0.00
14 T	Allyl chloride	0.941	0.972	-3.3	88	0.00
15 T	Acrylonitrile	0.335	0.348	-3.9	89	0.00
16 T	Acetone	0.394	0.393	0.3	88	0.00
17 T	Carbon Disulfide	1.188	1.174	1.2	86	0.00
18 T	Methyl Acetate	0.917	0.929	-1.3	89	0.00
19 T	Methyl tert-butyl Ether	2.092	2.100	-0.4	88	0.00
20 T	Methylene Chloride	0.668	0.628	6.0	89	0.00
21 T	trans-1,2-Dichloroethene	0.595	0.585	1.7	87	0.00
22 T	Diisopropyl ether	2.013	2.018	-0.2	88	0.00
23 T	Vinyl Acetate	1.722	1.815	-5.4	88	0.00
24 P	1,1-Dichloroethane	1.161	1.172	-0.9	88	0.00
25 T	2-Butanone	0.508	0.532	-4.7	89	0.00
26 T	2,2-Dichloropropane	1.049	1.098	-4.7	91	0.00
27 T	cis-1,2-Dichloroethene	0.735	0.725	1.4	87	0.00
28 T	Bromochloromethane	0.511	0.480	6.1	101	0.00
29 T	Tetrahydrofuran	0.312	0.317	-1.6	89	0.00
30 C	Chloroform	1.249	1.246	0.2#	90	0.00
31 T	Cyclohexane	0.956	0.980	-2.5	89	0.00
32 T	1,1,1-Trichloroethane	1.077	1.106	-2.7	88	0.00
33 S	1,2-Dichloroethane-d4	0.858	0.861	-0.3	108	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.357	0.375	-5.0	109	0.00
36 T	1,1-Dichloropropene	0.445	0.474	-6.5	91	0.00
37 T	Ethyl Acetate	0.553	0.582	-5.2	90	0.00
38 T	Carbon Tetrachloride	0.500	0.518	-3.6	88	0.00
39 T	Methylcyclohexane	0.578	0.619	-7.1	89	0.00
40 TM	Benzene	1.387	1.394	-0.5	88	0.00
41 T	Methacrylonitrile	0.283	0.305	-7.8	88	0.00
42 TM	1,2-Dichloroethane	0.557	0.562	-0.9	88	0.00
43 T	Isopropyl Acetate	0.880	0.905	-2.8	88	0.00
44 TM	Trichloroethene	0.393	0.350	10.9	89	0.00
45 C	1,2-Dichloropropane	0.340	0.325	4.4#	83	0.00
46 T	Dibromomethane	0.269	0.272	-1.1	88	0.00
47 T	Bromodichloromethane	0.482	0.507	-5.2	88	0.00
48 T	Methyl methacrylate	0.422	0.441	-4.5	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
 Data File : VX043932.D
 Acq On : 21 Nov 2024 13:57
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX112124

Quant Time: Nov 22 03:54:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	79	0.00
50 S	Toluene-d8	1.232	1.262	-2.4	108	0.00
51 T	4-Methyl-2-Pentanone	0.537	0.570	-6.1	89	0.00
52 CM	Toluene	0.830	0.855	-3.0#	88	0.00
53 T	t-1,3-Dichloropropene	0.491	0.522	-6.3	87	0.00
54 T	cis-1,3-Dichloropropene	0.539	0.562	-4.3	87	0.00
55 T	1,1,2-Trichloroethane	0.343	0.343	0.0	88	0.00
56 T	Ethyl methacrylate	0.534	0.567	-6.2	88	0.00
57 T	1,3-Dichloropropane	0.577	0.588	-1.9	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.273	0.275	-0.7	85	0.00
59 T	2-Hexanone	0.403	0.441	-9.4	89	0.00
60 T	Dibromochloromethane	0.344	0.361	-4.9	87	0.00
61 T	1,2-Dibromoethane	0.346	0.363	-4.9	89	0.00
62 S	4-Bromofluorobenzene	0.419	0.453	-8.1	111	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.330	0.332	-0.6	89	0.00
65 PM	Chlorobenzene	1.098	1.056	3.8	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.356	0.374	-5.1	89	0.00
67 C	Ethyl Benzene	1.861	1.898	-2.0#	88	0.00
68 T	m/p-Xylenes	0.694	0.712	-2.6	87	0.00
69 T	o-Xylene	0.678	0.696	-2.7	86	0.00
70 T	Styrene	1.121	1.159	-3.4	86	0.00
71 P	Bromoform	0.240	0.256	-6.7	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.843	3.784	1.5	89	0.00
74 T	N-amyl acetate	1.780	1.831	-2.9	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.316	1.299	1.3	90	0.00
76 T	1,2,3-Trichloropropane	1.096	1.096	0.0	89	0.00
77 T	Bromobenzene	0.917	0.882	3.8	87	0.00
78 T	n-propylbenzene	4.293	4.349	-1.3	88	0.00
79 T	2-Chlorotoluene	2.714	2.581	4.9	87	0.00
80 T	1,3,5-Trimethylbenzene	3.162	3.171	-0.3	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.402	0.400	0.5	87	0.00
82 T	4-Chlorotoluene	3.064	2.982	2.7	86	0.00
83 T	tert-Butylbenzene	3.134	3.144	-0.3	88	0.00
84 T	1,2,4-Trimethylbenzene	3.148	3.174	-0.8	89	0.00
85 T	sec-Butylbenzene	3.840	3.864	-0.6	87	0.00
86 T	p-Isopropyltoluene	3.119	3.213	-3.0	88	0.00
87 T	1,3-Dichlorobenzene	1.670	1.587	5.0	87	0.00
88 T	1,4-Dichlorobenzene	1.702	1.611	5.3	87	0.00
89 T	n-Butylbenzene	2.750	2.861	-4.0	86	0.00
90 T	Hexachloroethane	0.530	0.531	-0.2	87	0.00
91 T	1,2-Dichlorobenzene	1.685	1.606	4.7	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.264	0.269	-1.9	91	0.00
93 T	1,2,4-Trichlorobenzene	0.994	0.983	1.1	87	0.00
94 T	Hexachlorobutadiene	0.393	0.393	0.0	92	0.00
95 T	Naphthalene	3.576	3.707	-3.7	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112124\
Data File : VX043932.D
Acq On : 21 Nov 2024 13:57
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX112124

Quant Time: Nov 22 03:54:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.006	1.031	-2.5	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6