

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: EAEN05
Lab Code: CHEM Case No.: P4846 SAS No.: P4846 SDG NO.: P4846
Lab File ID: VX043950.D Date Analyzed: 11/22/2024
Instrument ID: MSVOA_X Time Analyzed: 11:22
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	119844	5.54	210147	6.75	179277	10.06
UPPER LIMIT	239688	6.038	420294	7.251	358554	10.555
LOWER LIMIT	59922	5.038	105074	6.251	89638.5	9.555
EPA SAMPLE NO.						
FSND-MW-EVAL-03D-20241113	109195	5.55	212971	6.76	181858	10.06
FSND-MW-EVAL-03S-20241113	111846	5.54	219449	6.76	192930	10.06
FSND-FD-3-20241113	107821	5.55	209572	6.76	187307	10.06
FSND-MW-EVAL-04S-20241113	112636	5.54	215010	6.76	191411	10.06
FSND-MW-26-20241113	110127	5.55	211509	6.76	181808	10.05
TB-1-20241113	108190	5.55	209028	6.76	182100	10.06
FSND-RB-3-20241113	111620	5.54	218890	6.76	191816	10.05
FSND-MW-24-20241113	108986	5.54	210122	6.76	184751	10.05
FSND-MW-15-20241113	110220	5.54	212874	6.76	183450	10.05
FSND-MW-EVAL-02D-20241113	106324	5.55	206768	6.76	176467	10.06
FSND-MW-EVAL-02S-20241113	121293	5.55	231454	6.76	196600	10.05
FSND-FD-4-20241113	113237	5.54	219483	6.76	191952	10.06
FSND-MW-EVAL-01D-20241113	108343	5.54	213438	6.76	185232	10.05
FSND-MW-EVAL-01S-20241113	106735	5.54	209074	6.76	185900	10.06
VX1122WBL01	114313	5.54	223688	6.76	192949	10.06
VX1122WBS01	122957	5.54	229417	6.76	195535	10.06

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: EAEN05
 Lab Code: CHEM Case No.: P4846 SAS No.: P4846 SDG NO.: P4846
 Lab File ID: VX043950.D Date Analyzed: 11/22/2024
 Instrument ID: MSVOA_X Time Analyzed: 11:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	119844	5.54	210147	6.75	179277	10.06
UPPER LIMIT	239688	6.038	420294	7.251	358554	10.555
LOWER LIMIT	59922	5.038	105074	6.251	89638.5	9.555
EPA SAMPLE NO.						

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.