

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: EAEN05
Lab Code: CHEM Case No.: P4846 SAS No.: P4846 SDG NO.: P4846
Lab File ID: VX043950.D Date Analyzed: 11/22/2024
Instrument ID: MSVOA_X Time Analyzed: 11:22
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	88069	12.018				
UPPER LIMIT	176138	12.518				
LOWER LIMIT	44034.5	11.518				
EPA SAMPLE NO.						
FSND-MW-EVAL-03D-20241113	75465	12.02				
FSND-MW-EVAL-03S-20241113	76810	12.02				
FSND-FD-3-20241113	74961	12.02				
FSND-MW-EVAL-04S-20241113	79010	12.02				
FSND-MW-26-20241113	70163	12.02				
TB-1-20241113	77306	12.02				
FSND-RB-3-20241113	77638	12.02				
FSND-MW-24-20241113	77269	12.02				
FSND-MW-15-20241113	74078	12.02				
FSND-MW-EVAL-02D-20241113	69943	12.02				
FSND-MW-EVAL-02S-20241113	81016	12.02				
FSND-FD-4-20241113	75953	12.02				
FSND-MW-EVAL-01D-20241113	77833	12.02				
FSND-MW-EVAL-01S-20241113	78265	12.02				
VX1122WBL01	79805	12.02				
VX1122WBS01	86659	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: EAEN05
 Lab Code: CHEM Case No.: P4846 SAS No.: P4846 SDG NO.: P4846
 Lab File ID: VX043950.D Date Analyzed: 11/22/2024
 Instrument ID: MSVOA_X Time Analyzed: 11:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	88069	12.018				
UPPER LIMIT	176138	12.518				
LOWER LIMIT	44034.5	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.