

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: EAEN05
 Lab Code: CHEM Case No.: P4846 SAS No.: P4846 SDG NO.: P4846
 Lab File ID: VX043977.D Date Analyzed: 11/25/2024
 Instrument ID: MSVOA_X Time Analyzed: 11:27
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	94886	12.018				
UPPER LIMIT	189772	12.518				
LOWER LIMIT	47443	11.518				
EPA SAMPLE NO.						
FSND-MW-EVAL-03D-20241113DL	73448	12.02				
FSND-MW-EVAL-04D-20241113	75257	12.02				
FSND-MW-EVAL-04D-20241113MS	88073	12.02				
FSND-MW-EVAL-04D-20241113MSD	91386	12.02				
FSND-MW-EVAL-01D-20241113DL	82992	12.02				
VX1125WBL01	90683	12.02				
VX1125WBS01	89446	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.