

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: P4852

Project: Webster School

RunNo.: LB133468

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/15/2024
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/15/2024
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/15/2024
Sample ID: CCB3 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/15/2024

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: P4852

Project: Webster School

RunNo.: LB133508

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	10/16/2024
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	10/16/2024
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	10/16/2024
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	10/16/2024
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	10/16/2024
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	10/16/2024
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	10/16/2024
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	11/15/2024
Chloride	mg/L	0.074	0.3000	J	0.011	0.6	11/15/2024
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	11/15/2024
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	11/15/2024
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	11/15/2024
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	11/15/2024
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	11/15/2024
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	11/15/2024
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	11/15/2024
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	11/15/2024
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	11/15/2024
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	11/15/2024
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	11/15/2024
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	11/15/2024

Initial and Continuing Calibration Blank Summary

Client: Kleinfelder

SDG No.: P4852

Project: Webster School

RunNo.: LB133520

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	11/19/2024
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	11/19/2024
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	11/19/2024
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	11/19/2024