

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: P4852

Project: Webster School

Run No.: LB133508

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133508BSS							
Bromide	mg/Kg	200	210		105	1	90-110	11/15/2024
Chloride	mg/Kg	60	62.7		104	1	90-110	11/15/2024
Fluoride	mg/Kg	40	42.2		106	1	90-110	11/15/2024
Nitrite	mg/Kg	60	62.9		105	1	90-110	11/15/2024
Nitrate	mg/Kg	50	52.5		105	1	90-110	11/15/2024
Sulfate	mg/Kg	300	310		103	1	90-110	11/15/2024
Orthophosphate as P	mg/Kg	100	104		104	1	90-110	11/15/2024

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: P4852

Project: Webster School

Run No.: LB133468

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB164961BS							
Hexavalent Chromium	mg/Kg	20	20.0		100	1	84-110	11/15/2024

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: P4852

Project: Webster School

Run No.: LB133520

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165004BS							
Ammonia as N	mg/Kg	50	51.0		102	1	90-110	11/19/2024