

### CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P4852 SAS No.: P4852 SDG NO.: P4852

Continuing Calib Date: 11/18/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 17:53 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.78       | 6.78      | 6.68      | 6.88  | 0.00       |
| Aroclor-1254-2       | (2) | 7.00       | 7.00      | 6.90      | 7.10  | 0.01       |
| Aroclor-1254-3       | (3) | 7.36       | 7.36      | 7.26      | 7.46  | 0.00       |
| Aroclor-1254-4       | (4) | 7.64       | 7.64      | 7.54      | 7.74  | 0.00       |
| Aroclor-1254-5       | (5) | 8.06       | 8.06      | 7.96      | 8.16  | 0.00       |
| Tetrachloro-m-xylene |     | 4.76       | 4.76      | 4.66      | 4.86  | 0.00       |
| Decachlorobiphenyl   |     | 10.68      | 10.67     | 10.57     | 10.77 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P4852 SAS No.: P4852 SDG NO.: P4852

Continuing Calib Date: 11/18/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 17:53 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.97       | 5.97      | 5.87      | 6.07 | 0.00       |
| Aroclor-1254-2 (2)   | 6.12       | 6.12      | 6.02      | 6.22 | 0.00       |
| Aroclor-1254-3 (3)   | 6.53       | 6.53      | 6.43      | 6.63 | 0.00       |
| Aroclor-1254-4 (4)   | 6.76       | 6.76      | 6.66      | 6.86 | 0.00       |
| Aroclor-1254-5 (5)   | 7.18       | 7.18      | 7.08      | 7.28 | 0.00       |
| Tetrachloro-m-xylene | 4.05       | 4.05      | 3.95      | 4.15 | 0.00       |
| Decachlorobiphenyl   | 9.21       | 9.21      | 9.11      | 9.31 | 0.00       |