

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM **Case No.:** P4852 **SAS No.:** P4852 **SDG NO.:** P4852

Continuing Calib Date: 11/19/2024 **Initial Calibration Date(s):** 11/08/2024 11/08/2024

Continuing Calib Time: 06:40 **Initial Calibration Time(s):** 02:45 10:39

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	4.76	4.76	4.66	4.86	0.00
Decachlorobiphenyl	10.68	10.67	10.57	10.77	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM **Case No.:** P4852 **SAS No.:** P4852 **SDG NO.:** P4852

Continuing Calib Date: 11/19/2024 **Initial Calibration Date(s):** 11/08/2024 11/08/2024

Continuing Calib Time: 06:40 **Initial Calibration Time(s):** 02:45 10:39

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl	9.21	9.21	9.11	9.31	0.00