

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P4852 SAS No.: P4852 SDG NO.: P4852

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 11/18/2024

Lab Sample No.: AR1660CCC500 Data File : PP068530.D Time Analyzed: 09:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.927	5.821	6.021	468.490	500.000	-6.3
Aroclor-1016-2	5.949	5.844	6.044	470.180	500.000	-6.0
Aroclor-1016-3	6.013	5.908	6.108	476.220	500.000	-4.8
Aroclor-1016-4	6.111	6.005	6.205	491.500	500.000	-1.7
Aroclor-1016-5	6.405	6.299	6.499	468.790	500.000	-6.2
Aroclor-1260-1	7.529	7.423	7.623	463.130	500.000	-7.4
Aroclor-1260-2	7.782	7.676	7.876	439.830	500.000	-12.0
Aroclor-1260-3	8.143	8.038	8.238	428.160	500.000	-14.4
Aroclor-1260-4	8.382	8.276	8.476	429.470	500.000	-14.1
Aroclor-1260-5	8.720	8.613	8.813	432.320	500.000	-13.5
Decachlorobiphenyl	10.679	10.572	10.772	44.200	50.000	-11.6
Tetrachloro-m-xylene	4.763	4.657	4.857	47.700	50.000	-4.6

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P4852 SAS No.: P4852 SDG NO.: P4852

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 11/18/2024

Lab Sample No.: AR1660CCC500 Data File : PP068530.D Time Analyzed: 09:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.157	5.058	5.258	484.470	500.000	-3.1
Aroclor-1016-2	5.178	5.079	5.279	473.470	500.000	-5.3
Aroclor-1016-3	5.358	5.259	5.459	456.240	500.000	-8.8
Aroclor-1016-4	5.397	5.298	5.498	465.350	500.000	-6.9
Aroclor-1016-5	5.615	5.517	5.717	467.310	500.000	-6.5
Aroclor-1260-1	6.659	6.560	6.760	459.160	500.000	-8.2
Aroclor-1260-2	6.845	6.747	6.947	456.200	500.000	-8.8
Aroclor-1260-3	7.003	6.905	7.105	462.240	500.000	-7.6
Aroclor-1260-4	7.477	7.378	7.578	472.290	500.000	-5.5
Aroclor-1260-5	7.715	7.617	7.817	461.250	500.000	-7.8
Decachlorobiphenyl	9.208	9.110	9.310	45.210	50.000	-9.6
Tetrachloro-m-xylene	4.052	3.951	4.151	49.210	50.000	-1.6