

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P4852 SAS No.: P4852 SDG NO.: P4852

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/18/2024

Lab Sample No.: AR1660CCC500 Data File : PP068542.D Time Analyzed: 13:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.926	5.821	6.021	457.730	500.000	-8.5
Aroclor-1016-2	5.948	5.844	6.044	458.850	500.000	-8.2
Aroclor-1016-3	6.011	5.908	6.108	455.320	500.000	-8.9
Aroclor-1016-4	6.109	6.005	6.205	487.870	500.000	-2.4
Aroclor-1016-5	6.403	6.299	6.499	464.930	500.000	-7.0
Aroclor-1260-1	7.528	7.423	7.623	435.100	500.000	-13.0
Aroclor-1260-2	7.781	7.676	7.876	434.590	500.000	-13.1
Aroclor-1260-3	8.142	8.038	8.238	432.090	500.000	-13.6
Aroclor-1260-4	8.380	8.276	8.476	420.380	500.000	-15.9
Aroclor-1260-5	8.718	8.613	8.813	426.480	500.000	-14.7
Decachlorobiphenyl	10.678	10.572	10.772	44.870	50.000	-10.3
Tetrachloro-m-xylene	4.761	4.657	4.857	48.070	50.000	-3.9

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P4852 SAS No.: P4852 SDG NO.: P4852

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/18/2024

Lab Sample No.: AR1660CCC500 Data File : PP068542.D Time Analyzed: 13:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.157	5.058	5.258	510.770	500.000	2.2
Aroclor-1016-2	5.177	5.079	5.279	500.810	500.000	0.2
Aroclor-1016-3	5.357	5.259	5.459	486.900	500.000	-2.6
Aroclor-1016-4	5.397	5.298	5.498	489.450	500.000	-2.1
Aroclor-1016-5	5.616	5.517	5.717	479.630	500.000	-4.1
Aroclor-1260-1	6.659	6.560	6.760	448.840	500.000	-10.2
Aroclor-1260-2	6.845	6.747	6.947	444.970	500.000	-11.0
Aroclor-1260-3	7.003	6.905	7.105	450.050	500.000	-10.0
Aroclor-1260-4	7.477	7.378	7.578	459.060	500.000	-8.2
Aroclor-1260-5	7.716	7.617	7.817	444.870	500.000	-11.0
Decachlorobiphenyl	9.208	9.110	9.310	45.610	50.000	-8.8
Tetrachloro-m-xylene	4.051	3.951	4.151	50.540	50.000	1.1