

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P4852 SAS No.: P4852 SDG NO.: P4852

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL03 Date Analyzed: 11/18/2024

Lab Sample No.: AR1660CCC500 Data File : PP068554.D Time Analyzed: 17:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.925	5.821	6.021	464.840	500.000	-7.0
Aroclor-1016-2	5.947	5.844	6.044	474.020	500.000	-5.2
Aroclor-1016-3	6.011	5.908	6.108	474.730	500.000	-5.1
Aroclor-1016-4	6.109	6.005	6.205	481.140	500.000	-3.8
Aroclor-1016-5	6.403	6.299	6.499	460.140	500.000	-8.0
Aroclor-1260-1	7.527	7.423	7.623	432.380	500.000	-13.5
Aroclor-1260-2	7.780	7.676	7.876	424.720	500.000	-15.1
Aroclor-1260-3	8.141	8.038	8.238	409.620	500.000	-18.1
Aroclor-1260-4	8.380	8.276	8.476	418.660	500.000	-16.3
Aroclor-1260-5	8.717	8.613	8.813	414.130	500.000	-17.2
Decachlorobiphenyl	10.676	10.572	10.772	43.350	50.000	-13.3
Tetrachloro-m-xylene	4.761	4.657	4.857	49.690	50.000	-0.6

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P4852 SAS No.: P4852 SDG NO.: P4852

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL03 Date Analyzed: 11/18/2024

Lab Sample No.: AR1660CCC500 Data File : PP068554.D Time Analyzed: 17:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.157	5.058	5.258	495.490	500.000	-0.9
Aroclor-1016-2	5.178	5.079	5.279	481.120	500.000	-3.8
Aroclor-1016-3	5.358	5.259	5.459	464.580	500.000	-7.1
Aroclor-1016-4	5.397	5.298	5.498	469.540	500.000	-6.1
Aroclor-1016-5	5.616	5.517	5.717	492.550	500.000	-1.5
Aroclor-1260-1	6.659	6.560	6.760	456.050	500.000	-8.8
Aroclor-1260-2	6.846	6.747	6.947	459.320	500.000	-8.1
Aroclor-1260-3	7.003	6.905	7.105	463.440	500.000	-7.3
Aroclor-1260-4	7.477	7.378	7.578	456.610	500.000	-8.7
Aroclor-1260-5	7.716	7.617	7.817	449.570	500.000	-10.1
Decachlorobiphenyl	9.209	9.110	9.310	44.370	50.000	-11.3
Tetrachloro-m-xylene	4.051	3.951	4.151	49.520	50.000	-1.0