

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P4852 SAS No.: P4852 SDG NO.: P4852

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL04 Date Analyzed: 11/19/2024

Lab Sample No.: AR1660CCC500 Data File : PP068602.D Time Analyzed: 11:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.923	5.821	6.021	462.180	500.000	-7.6
Aroclor-1016-2	5.945	5.844	6.044	481.740	500.000	-3.7
Aroclor-1016-3	6.008	5.908	6.108	490.180	500.000	-2.0
Aroclor-1016-4	6.106	6.005	6.205	490.380	500.000	-1.9
Aroclor-1016-5	6.400	6.299	6.499	462.630	500.000	-7.5
Aroclor-1260-1	7.525	7.423	7.623	477.630	500.000	-4.5
Aroclor-1260-2	7.777	7.676	7.876	459.180	500.000	-8.2
Aroclor-1260-3	8.139	8.038	8.238	431.860	500.000	-13.6
Aroclor-1260-4	8.376	8.276	8.476	426.790	500.000	-14.6
Aroclor-1260-5	8.715	8.613	8.813	422.590	500.000	-15.5
Decachlorobiphenyl	10.674	10.572	10.772	44.450	50.000	-11.1
Tetrachloro-m-xylene	4.758	4.657	4.857	49.130	50.000	-1.7

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: P4852 SAS No.: P4852 SDG NO.: P4852

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL04 Date Analyzed: 11/19/2024

Lab Sample No.: AR1660CCC500 Data File : PP068602.D Time Analyzed: 11:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.154	5.058	5.258	509.510	500.000	1.9
Aroclor-1016-2	5.174	5.079	5.279	473.120	500.000	-5.4
Aroclor-1016-3	5.355	5.259	5.459	439.650	500.000	-12.1
Aroclor-1016-4	5.394	5.298	5.498	442.810	500.000	-11.4
Aroclor-1016-5	5.614	5.517	5.717	493.170	500.000	-1.4
Aroclor-1260-1	6.657	6.560	6.760	494.420	500.000	-1.1
Aroclor-1260-2	6.843	6.747	6.947	469.380	500.000	-6.1
Aroclor-1260-3	7.001	6.905	7.105	467.400	500.000	-6.5
Aroclor-1260-4	7.475	7.378	7.578	472.750	500.000	-5.5
Aroclor-1260-5	7.713	7.617	7.817	458.570	500.000	-8.3
Decachlorobiphenyl	9.205	9.110	9.310	45.250	50.000	-9.5
Tetrachloro-m-xylene	4.049	3.951	4.151	49.290	50.000	-1.4