

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP110724

Review By	yogesh	Review On	11/11/2024 9:23:03 AM
Supervise By	Ankita	Supervise On	11/11/2024 9:48:52 AM
SubDirectory	PP110724	HP Acquire Method	HP Processing Method PP110724
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP068359.D	08 Nov 2024 02:13	YP\AJ	Ok
2	I.BLK	PP068360.D	08 Nov 2024 02:29	YP\AJ	Ok
3	AR1660ICC1000	PP068361.D	08 Nov 2024 02:45	YP\AJ	Ok,M
4	AR1660ICC750	PP068362.D	08 Nov 2024 03:02	YP\AJ	Ok
5	AR1660ICC500	PP068363.D	08 Nov 2024 03:18	YP\AJ	Ok
6	AR1660ICC250	PP068364.D	08 Nov 2024 03:34	YP\AJ	Ok,M
7	AR1660ICC050	PP068365.D	08 Nov 2024 03:51	YP\AJ	Ok,M
8	AR1221ICC500	PP068366.D	08 Nov 2024 04:07	YP\AJ	Ok
9	AR1232ICC500	PP068367.D	08 Nov 2024 04:23	YP\AJ	Ok
10	AR1242ICC1000	PP068368.D	08 Nov 2024 04:40	YP\AJ	Ok
11	AR1242ICC750	PP068369.D	08 Nov 2024 04:56	YP\AJ	Ok,M
12	AR1242ICC500	PP068370.D	08 Nov 2024 05:12	YP\AJ	Ok
13	AR1242ICC250	PP068371.D	08 Nov 2024 05:29	YP\AJ	Ok,M
14	AR1242ICC050	PP068372.D	08 Nov 2024 05:45	YP\AJ	Ok
15	AR1248ICC1000	PP068373.D	08 Nov 2024 06:01	YP\AJ	Ok
16	AR1248ICC750	PP068374.D	08 Nov 2024 06:18	YP\AJ	Ok
17	AR1248ICC500	PP068375.D	08 Nov 2024 06:34	YP\AJ	Ok
18	AR1248ICC250	PP068376.D	08 Nov 2024 06:50	YP\AJ	Ok,M
19	AR1248ICC050	PP068377.D	08 Nov 2024 07:07	YP\AJ	Ok
20	AR1254ICC1000	PP068378.D	08 Nov 2024 07:23	YP\AJ	Ok,M
21	AR1254ICC750	PP068379.D	08 Nov 2024 07:39	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP110724

Review By	yogesh	Review On	11/11/2024 9:23:03 AM
Supervise By	Ankita	Supervise On	11/11/2024 9:48:52 AM
SubDirectory	PP110724	HP Acquire Method	HP Processing Method PP110724
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP068380.D	08 Nov 2024 07:56	YP\AJ	Ok,M
23	AR1254ICC250	PP068381.D	08 Nov 2024 08:12	YP\AJ	Ok,M
24	AR1254ICC050	PP068382.D	08 Nov 2024 08:28	YP\AJ	Ok,M
25	AR1262ICC500	PP068383.D	08 Nov 2024 08:45	YP\AJ	Ok
26	AR1268ICC1000	PP068384.D	08 Nov 2024 09:01	YP\AJ	Ok,M
27	AR1268ICC750	PP068385.D	08 Nov 2024 09:17	YP\AJ	Ok,M
28	AR1268ICC500	PP068386.D	08 Nov 2024 09:34	YP\AJ	Ok
29	AR1268ICC250	PP068387.D	08 Nov 2024 09:50	YP\AJ	Ok,M
30	AR1268ICC050	PP068388.D	08 Nov 2024 10:39	YP\AJ	Ok,M
31	PP110724ICV500	PP068389.D	08 Nov 2024 10:55	YP\AJ	Ok,M
32	AR1242ICV500	PP068390.D	08 Nov 2024 11:12	YP\AJ	Ok,M
33	AR1248ICV500	PP068391.D	08 Nov 2024 11:28	YP\AJ	Ok,M
34	AR1254ICV500	PP068392.D	08 Nov 2024 11:45	YP\AJ	Ok,M
35	AR1268ICV500	PP068393.D	08 Nov 2024 12:01	YP\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111824

Review By	yogesh	Review On	11/19/2024 9:08:31 AM
Supervise By	Ankita	Supervise On	11/19/2024 11:25:02 AM
SubDirectory	PP111824	HP Acquire Method	HP Processing Method PP110724
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP068529.D	18 Nov 2024 09:22	YP\AJ	Ok
2	AR1660CCC500	PP068530.D	18 Nov 2024 09:38	YP\AJ	Ok
3	I.BLK	PP068531.D	18 Nov 2024 09:54	YP\AJ	Ok
4	P4867-02RE	PP068532.D	18 Nov 2024 10:11	YP\AJ	Confirms
5	P4867-03RE	PP068533.D	18 Nov 2024 10:27	YP\AJ	Confirms
6	P4857-01	PP068534.D	18 Nov 2024 10:43	YP\AJ	Ok,M
7	P4859-01	PP068535.D	18 Nov 2024 11:00	YP\AJ	Ok,M
8	P4859-01MS	PP068536.D	18 Nov 2024 11:16	YP\AJ	Ok,M
9	P4859-01MSD	PP068537.D	18 Nov 2024 11:33	YP\AJ	Ok,M
10	PB164991BL	PP068538.D	18 Nov 2024 11:49	YP\AJ	Ok
11	PB164991BS	PP068539.D	18 Nov 2024 12:05	YP\AJ	Ok
12	P4870-01	PP068540.D	18 Nov 2024 12:22	YP\AJ	Ok
13	P4870-04	PP068541.D	18 Nov 2024 12:38	YP\AJ	Ok
14	AR1660CCC500	PP068542.D	18 Nov 2024 13:05	YP\AJ	Ok
15	I.BLK	PP068543.D	18 Nov 2024 13:21	YP\AJ	Ok
16	P4848-01	PP068544.D	18 Nov 2024 13:38	YP\AJ	Ok
17	P4849-01	PP068545.D	18 Nov 2024 13:54	YP\AJ	Ok,M
18	P4850-01	PP068546.D	18 Nov 2024 14:10	YP\AJ	Ok
19	P4851-01	PP068547.D	18 Nov 2024 14:27	YP\AJ	Ok
20	P4852-01	PP068548.D	18 Nov 2024 14:43	YP\AJ	Ok
21	P4852-02	PP068549.D	18 Nov 2024 15:00	YP\AJ	Not Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111824

Review By	yogesh	Review On	11/19/2024 9:08:31 AM
Supervise By	Ankita	Supervise On	11/19/2024 11:25:02 AM
SubDirectory	PP111824	HP Acquire Method	HP Processing Method PP110724
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	P4852-03	PP068550.D	18 Nov 2024 15:16	YP\AJ	Ok
23	P4857-03	PP068551.D	18 Nov 2024 15:32	YP\AJ	Ok,M
24	P4870-07	PP068552.D	18 Nov 2024 15:49	YP\AJ	Ok
25	P4870-10	PP068553.D	18 Nov 2024 16:05	YP\AJ	Ok
26	AR1660CCC500	PP068554.D	18 Nov 2024 17:04	YP\AJ	Ok
27	AR1242CCC500	PP068555.D	18 Nov 2024 17:20	YP\AJ	Ok
28	AR1248CCC500	PP068556.D	18 Nov 2024 17:37	YP\AJ	Ok
29	AR1254CCC500	PP068557.D	18 Nov 2024 17:53	YP\AJ	Ok,M
30	I.BLK	PP068558.D	18 Nov 2024 18:09	YP\AJ	Ok
31	PB165027BL	PP068559.D	18 Nov 2024 18:26	YP\AJ	Ok
32	PB165027BS	PP068560.D	18 Nov 2024 18:42	YP\AJ	Ok
33	P4862-01	PP068561.D	18 Nov 2024 18:58	YP\AJ	Ok,M
34	P4862-02	PP068562.D	18 Nov 2024 19:15	YP\AJ	Ok
35	P4862-03	PP068563.D	18 Nov 2024 19:31	YP\AJ	Ok
36	P4862-04	PP068564.D	18 Nov 2024 19:47	YP\AJ	Ok
37	P4862-05	PP068565.D	18 Nov 2024 20:04	YP\AJ	Ok
38	P4862-06	PP068566.D	18 Nov 2024 20:20	YP\AJ	Ok,M
39	P4862-07	PP068567.D	18 Nov 2024 20:36	YP\AJ	Ok,M
40	P4862-08	PP068568.D	18 Nov 2024 20:53	YP\AJ	Ok,M
41	AR1660CCC500	PP068569.D	19 Nov 2024 01:25	YP\AJ	Ok
42	AR1242CCC500	PP068570.D	19 Nov 2024 01:57	YP\AJ	Ok
43	AR1248CCC500	PP068571.D	19 Nov 2024 02:14	YP\AJ	Ok
44	AR1254CCC500	PP068572.D	19 Nov 2024 02:30	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111824

Review By	yogesh	Review On	11/19/2024 9:08:31 AM
Supervise By	Ankita	Supervise On	11/19/2024 11:25:02 AM
SubDirectory	PP111824	HP Acquire Method	HP Processing Method PP110724
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	I.BLK	PP068573.D	19 Nov 2024 02:46	YP\AJ	Ok,M
46	P4865-01	PP068574.D	18 Nov 2024 23:14	YP\AJ	Ok,M
47	P4865-02	PP068575.D	18 Nov 2024 23:31	YP\AJ	Ok
48	P4865-03	PP068576.D	19 Nov 2024 03:35	YP\AJ	Ok,M
49	P4865-04	PP068577.D	19 Nov 2024 03:52	YP\AJ	Ok,M
50	P4890-03	PP068578.D	19 Nov 2024 04:08	YP\AJ	Ok,M
51	PB165038BL	PP068579.D	19 Nov 2024 04:24	YP\AJ	Ok
52	PB165038BS	PP068580.D	19 Nov 2024 04:41	YP\AJ	Ok,M
53	PB165038BSD	PP068581.D	19 Nov 2024 04:57	YP\AJ	Ok,M
54	P4864-01	PP068582.D	19 Nov 2024 05:13	YP\AJ	Ok
55	P4890-02	PP068583.D	19 Nov 2024 05:29	YP\AJ	ReRun
56	AR1660CCC500	PP068584.D	19 Nov 2024 06:07	YP\AJ	Ok
57	AR1242CCC500	PP068585.D	19 Nov 2024 06:23	YP\AJ	Ok
58	AR1248CCC500	PP068586.D	19 Nov 2024 06:40	YP\AJ	Ok
59	AR1254CCC500	PP068587.D	19 Nov 2024 06:56	YP\AJ	Ok,M
60	I.BLK	PP068588.D	19 Nov 2024 07:13	YP\AJ	Ok
61	PB165043BL	PP068589.D	19 Nov 2024 07:29	YP\AJ	Ok
62	P4890-01	PP068590.D	19 Nov 2024 07:45	YP\AJ	Not Ok
63	P4890-06	PP068591.D	19 Nov 2024 08:02	YP\AJ	Ok,M
64	P4887-01	PP068592.D	19 Nov 2024 08:18	YP\AJ	Ok
65	P4887-05	PP068593.D	19 Nov 2024 08:34	YP\AJ	Ok,M
66	P4889-01	PP068594.D	19 Nov 2024 08:51	YP\AJ	Ok,M
67	P4889-03	PP068595.D	19 Nov 2024 09:23	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111824

Review By	yogesh	Review On	11/19/2024 9:08:31 AM
Supervise By	Ankita	Supervise On	11/19/2024 11:25:02 AM
SubDirectory	PP111824	HP Acquire Method	HP Processing Method PP110724
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

68	P4889-05	PP068596.D	19 Nov 2024 09:39	YP\AJ	Ok,M
69	P4869-01	PP068597.D	19 Nov 2024 09:56	YP\AJ	Ok,M
70	AR1660CCC500	PP068598.D	19 Nov 2024 10:44	YP\AJ	Ok,M
71	AR1242CCC500	PP068599.D	19 Nov 2024 11:00	YP\AJ	Ok
72	AR1254CCC500	PP068600.D	19 Nov 2024 11:17	YP\AJ	Ok,M
73	I.BLK	PP068601.D	19 Nov 2024 11:33	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111924

Review By	yogesh	Review On	11/20/2024 8:01:50 AM
Supervise By	Ankita	Supervise On	11/20/2024 12:02:39 PM
SubDirectory	PP111924	HP Acquire Method	HP Processing Method PP110724
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	AR1660CCC500	PP068602.D	19 Nov 2024 11:50	YP\AJ	Ok,M
2	AR1242CCC500	PP068603.D	19 Nov 2024 12:06	YP\AJ	Ok
3	AR1248CCC500	PP068604.D	19 Nov 2024 12:22	YP\AJ	Ok
4	AR1254CCC500	PP068605.D	19 Nov 2024 12:39	YP\AJ	Ok,M
5	I.BLK	PP068606.D	19 Nov 2024 12:55	YP\AJ	Ok
6	P4852-02	PP068607.D	19 Nov 2024 13:11	YP\AJ	Ok
7	P4890-02RE	PP068608.D	19 Nov 2024 13:28	YP\AJ	Confirms
8	PB165084BL	PP068609.D	19 Nov 2024 13:44	YP\AJ	Ok
9	PB165084BS	PP068610.D	19 Nov 2024 14:00	YP\AJ	Ok,M
10	P4888-01	PP068611.D	19 Nov 2024 14:17	YP\AJ	Ok
11	P4888-02	PP068612.D	19 Nov 2024 14:33	YP\AJ	Ok
12	P4890-01	PP068613.D	19 Nov 2024 14:51	YP\AJ	Ok,M
13	AR1660CCC500	PP068614.D	19 Nov 2024 15:29	YP\AJ	Ok,M
14	I.BLK	PP068615.D	19 Nov 2024 15:45	YP\AJ	Ok

M : Manual Integration