

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4852

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VD080061.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VD1118SBSD01	cis-1,2-Dichloroethene	20	20.3	ug/Kg	102	10		82	123	20
	1,1,1-Trichloroethane	20	20.6	ug/Kg	103	1		80	126	20
	Benzene	20	20.8	ug/Kg	104	7		84	121	20
	Trichloroethene	20	20.5	ug/Kg	103	5		83	122	20
	Toluene	20	21.3	ug/Kg	106	8		83	122	20
	Ethyl Benzene	20	19.3	ug/Kg	97	4		82	124	20
	m/p-Xylenes	40	39.5	ug/Kg	99	2		83	124	20
	o-Xylene	20	19.7	ug/Kg	99	0		83	123	20
	Isopropylbenzene	20	19.9	ug/Kg	100	7		82	124	20