

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4852
Client: Kleinfelder
Analytical Method: SW8260D **Datafile :** VY020375.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1121SBS01	cis-1,2-Dichloroethene	20	18.9	ug/Kg	95			82	123	
	1,1,1-Trichloroethane	20	20.2	ug/Kg	101			80	126	
	Benzene	20	19.6	ug/Kg	98			84	121	
	Trichloroethene	20	19.3	ug/Kg	97			83	122	
	Toluene	20	19.3	ug/Kg	97			83	122	
	Ethyl Benzene	20	19.0	ug/Kg	95			82	124	
	m/p-Xylenes	40	39.2	ug/Kg	98			83	124	
	o-Xylene	20	19.5	ug/Kg	98			83	123	
	Isopropylbenzene	20	18.4	ug/Kg	92			82	124	