

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
 Lab Code: CHEM Case No.: P4852 SAS No.: P4852 SDG NO.: P4852
 Lab File ID: VD080058.D Date Analyzed: 11/18/2024
 Instrument ID: MSVOA_D Time Analyzed: 09:51
 GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	138959	13.516				
UPPER LIMIT	277918	14.016				
LOWER LIMIT	69479.5	13.016				
EPA SAMPLE NO.						
COMP-2	10070 *	13.52				
VD1118SBL01	122131	13.52				
VD1118SBS01	141556	13.52				
VD1118SBSD01	143766	13.52				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.