

DATA PACKAGE

VOLATILE ORGANICS
GENERAL CHEMISTRY
METALS

PROJECT NAME : TRANSFER STATION-SPDES

TULLY ENVIRONMENTAL, INC

127-50 Northern Blvd.

Flushing, NY - 11368

Phone No: 718-446-7000

ORDER ID : P4853

ATTENTION : Dean Devoe



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) VOC-BTEX- Case Narrative	4
2.2) Metals-AES- Case Narrative	6
2.3) Genchem- Case Narrative	7
3) Qualifier Page	8
4) QA Checklist	10
5) VOC-BTEX Data	11
6) Metals-AES Data	18
7) Genchem Data	23
8) Shipping Document	27
8.1) CHAIN OF CUSTODY	28
8.2) Lab Certificate	29
8.3) Internal COC	30

1

2

3

4

5

6

7

8

Cover Page

Order ID : P4853

Project ID : Transfer Station-SPDES

Client : Tully Environmental, Inc

Lab Sample Number

P4853-01
P4853-02

Client Sample Number

001-WILLETS-PT-BLVD(NOV)
002-35TH-AVE(NOV)

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 11/26/2024

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Tully Environmental, Inc

Project Name: Transfer Station-SPDES

Project # N/A

Chemtech Project # P4853

Test Name: VOC-BTEX

A. Number of Samples and Date of Receipt:

2 Water samples were received on 11/14/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, BOD5, Metals Group 10, Oil and Grease, TSS and VOC-BTEX. This data package contains results for VOC-BTEX.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_N were done using GC column RXI-624SIL MS 30m 0.25mm 1.4 um. Cat#13868. The analysis of VOC-BTEX was based on method 624.1.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

Samples 001-WILLETS-PT-BLVD(NOV), 002-35TH-AVE(NOV) were diluted due to high concentrations.

E. Additional Comments:

Trip Blank was not provided with this set of samples.

As per method, MS/MSD is required to be performed with the sample analysis.

However, Lab did not receive sufficient volume to perform the MS/MSD therefore MS/MSD were not performed for this project. However, Lab has performed LCS/LCSD instead.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <35% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 35% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tully Environmental, Inc

Project Name: Transfer Station-SPDES

Project # N/A

Chemtech Project # P4853

Test Name: Metals Group 10

A. Number of Samples and Date of Receipt:

2 Water samples were received on 11/14/2024.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Ammonia, BOD5, Metals Group 10, Oil and Grease, TSS and VOC-BTEX. This data package contains results for Metals Group 10.

C. Analytical Techniques:

The analysis and digestion of Metals Group 10 was based on method 200.7.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (001-WILLETS-PT-BLVD(OCT)MS) analysis met criteria for all samples except for Iron due to Chemical Interference during Digestion Process.

The Matrix Spike Duplicate (001-WILLETS-PT-BLVD(OCT)MSD) analysis met criteria for all samples except for Iron due to Chemical Interference during Digestion Process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tully Environmental, Inc

Project Name: Transfer Station-SPDES

Project # N/A

Chemtech Project # P4853

Test Name: Oil and Grease,Ammonia,BOD5,TSS

A. Number of Samples and Date of Receipt:

2 Water samples were received on 11/14/2024.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Ammonia, BOD5, Metals Group 10, Oil and Grease, TSS and VOC-BTEX. This data package contains results for Oil and Grease,Ammonia,BOD5,TSS.

C. Analytical Techniques:

The analysis of Oil and Grease was based on method 1664A, The analysis of TSS was based on method SM2540 D, The analysis of Ammonia was based on method SM4500-NH3 and The analysis of BOD5 was based on method SM5210 B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

As per method 1664A, MS/MSD is required to be performed with the sample analysis. However, Lab did not receive sufficient volume to perform the MS/MSD for P4853 therefore Lab reported MS-MSD from P4900 and P4937.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: <ol style="list-style-type: none"> (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: P4853

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 11/26/2024

Hit Summary Sheet
SW-846

SDG No.: P4853
Client: Tully Environmental, Inc

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: P4853-01	001-WILLETS-PT-BLVD(NOV) 001-WILLETS-PT- Water	Toluene		220	E	0.72	5.00	ug/L
		Total Voc :		220				
		Total Concentration:		220				
Client ID: P4853-01DL	001-WILLETS-PT-BLVD(NOV)DL 001-WILLETS-PT- Water	Toluene		170	D	3.60	25.0	ug/L
		Total Voc :		170				
		Total Concentration:		170				
Client ID: P4853-02	002-35TH-AVE(NOV) 002-35TH-AVE(NC Water	Toluene		210	E	0.72	5.00	ug/L
		Total Voc :		210				
		Total Concentration:		210				
Client ID: P4853-02DL	002-35TH-AVE(NOV)DL 002-35TH-AVE(NC Water	Toluene		180	D	3.60	25.0	ug/L
		Total Voc :		180				
		Total Concentration:		180				



SAMPLE DATA

Report of Analysis

Client:	Tully Environmental, Inc		Date Collected:	11/13/24	
Project:	Transfer Station-SPDES		Date Received:	11/14/24	
Client Sample ID:	001-WILLETS-PT-BLVD(NOV)		SDG No.:	P4853	
Lab Sample ID:	P4853-01		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-BTEX	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084874.D	1		11/15/24 17:12	VN111524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	0.69	U	0.69	5.00	ug/L
108-88-3	Toluene	220	E	0.72	5.00	ug/L
100-41-4	Ethyl Benzene	0.73	U	0.73	5.00	ug/L
179601-23-1	m/p-Xylenes	1.70	U	1.70	10.0	ug/L
95-47-6	o-Xylene	0.82	U	0.82	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	31.3		91 - 110	104%	SPK: 30
2037-26-5	Toluene-d8	28.5		91 - 112	95%	SPK: 30
460-00-4	4-Bromofluorobenzene	25.4		63 - 112	85%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	22200	7.806			
540-36-3	1,4-Difluorobenzene	116000	9.1			
3114-55-4	Chlorobenzene-d5	108000	11.865			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tully Environmental, Inc		Date Collected:	11/13/24	
Project:	Transfer Station-SPDES		Date Received:	11/14/24	
Client Sample ID:	001-WILLETS-PT-BLVD(NOV)DL		SDG No.:	P4853	
Lab Sample ID:	P4853-01DL		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-BTEX	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084991.D	5		11/21/24 12:18	VN112124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	3.50	UD	3.50	25.0	ug/L
108-88-3	Toluene	170	D	3.60	25.0	ug/L
100-41-4	Ethyl Benzene	3.70	UD	3.70	25.0	ug/L
179601-23-1	m/p-Xylenes	8.60	UD	8.60	50.0	ug/L
95-47-6	o-Xylene	4.10	UD	4.10	25.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.8		91 - 110	103%	SPK: 30
2037-26-5	Toluene-d8	28.4		91 - 112	95%	SPK: 30
460-00-4	4-Bromofluorobenzene	25.8		63 - 112	86%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	29900	7.812			
540-36-3	1,4-Difluorobenzene	161000	9.094			
3114-55-4	Chlorobenzene-d5	139000	11.865			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tully Environmental, Inc		Date Collected:	11/13/24	
Project:	Transfer Station-SPDES		Date Received:	11/14/24	
Client Sample ID:	002-35TH-AVE(NOV)		SDG No.:	P4853	
Lab Sample ID:	P4853-02		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-BTEX	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084875.D	1		11/15/24 17:36	VN111524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	0.69	U	0.69	5.00	ug/L
108-88-3	Toluene	210	E	0.72	5.00	ug/L
100-41-4	Ethyl Benzene	0.73	U	0.73	5.00	ug/L
179601-23-1	m/p-Xylenes	1.70	U	1.70	10.0	ug/L
95-47-6	o-Xylene	0.82	U	0.82	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.8		91 - 110	103%	SPK: 30
2037-26-5	Toluene-d8	28.2		91 - 112	94%	SPK: 30
460-00-4	4-Bromofluorobenzene	26.4		63 - 112	88%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	23600	7.806			
540-36-3	1,4-Difluorobenzene	123000	9.1			
3114-55-4	Chlorobenzene-d5	115000	11.865			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tully Environmental, Inc		Date Collected:	11/13/24	
Project:	Transfer Station-SPDES		Date Received:	11/14/24	
Client Sample ID:	002-35TH-AVE(NOV)DL		SDG No.:	P4853	
Lab Sample ID:	P4853-02DL		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-BTEX	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084992.D	5		11/21/24 12:42	VN112124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
71-43-2	Benzene	3.50	UD	3.50	25.0	ug/L
108-88-3	Toluene	180	D	3.60	25.0	ug/L
100-41-4	Ethyl Benzene	3.70	UD	3.70	25.0	ug/L
179601-23-1	m/p-Xylenes	8.60	UD	8.60	50.0	ug/L
95-47-6	o-Xylene	4.10	UD	4.10	25.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.1		91 - 110	100%	SPK: 30
2037-26-5	Toluene-d8	28.0		91 - 112	93%	SPK: 30
460-00-4	4-Bromofluorobenzene	25.2		63 - 112	84%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	29500	7.812			
540-36-3	1,4-Difluorobenzene	147000	9.1			
3114-55-4	Chlorobenzene-d5	131000	11.865			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	P4853	OrderDate:	11/14/2024 11:45:00 AM
Client:	Tully Environmental, Inc	Project:	Transfer Station-SPDES
Contact:	Dean Devoe	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4853-01	001-WILLETS-PT-BLV D(NOV)	Water			11/13/24			11/14/24
			VOC-BTEX	624.1			11/15/24	
P4853-01DL	001-WILLETS-PT-BLV D(NOV)DL	Water			11/13/24			11/14/24
			VOC-BTEX	624.1			11/21/24	
P4853-02	002-35TH-AVE(NOV)	Water			11/13/24			11/14/24
			VOC-BTEX	624.1			11/15/24	
P4853-02DL	002-35TH-AVE(NOV)D L	Water			11/13/24			11/14/24
			VOC-BTEX	624.1			11/21/24	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: P4853 **Order ID:** P4853
Client: Tully Environmental, Inc **Project ID:** Transfer Station-SPDES

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : 001-WILLETS-PT-BLVD(NOV)								
P4853-01	001-WILLETS-PT-BLVD(NOV)	Water	Copper	21.7		1.52	10.0	ug/L
P4853-01	001-WILLETS-PT-BLVD(NOV)	Water	Iron	8910		12.4	50.0	ug/L
P4853-01	001-WILLETS-PT-BLVD(NOV)	Water	Lead	28.9		1.57	6.00	ug/L



SAMPLE DATA

Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	11/13/24
Project:	Transfer Station-SPDES	Date Received:	11/14/24
Client Sample ID:	001-WILLETS-PT-BLVD(NOV)	SDG No.:	P4853
Lab Sample ID:	P4853-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-50-8	Copper	21.7		1	1.52	10.0	ug/L	11/15/24 09:20	11/15/24 13:33	EPA 200.7	
7439-89-6	Iron	8910	N	1	12.4	50.0	ug/L	11/15/24 09:20	11/15/24 13:33	EPA 200.7	
7439-92-1	Lead	28.9		1	1.57	6.00	ug/L	11/15/24 09:20	11/15/24 13:33	EPA 200.7	

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group 10			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	11/13/24
Project:	Transfer Station-SPDES	Date Received:	11/14/24
Client Sample ID:	002-35TH-AVE(NOV)	SDG No.:	P4853
Lab Sample ID:	P4853-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-50-8	Copper	1.52	U	1	1.52	10.0	ug/L	11/15/24 09:20	11/15/24 13:38	EPA 200.7	
7439-89-6	Iron	12.4	UN	1	12.4	50.0	ug/L	11/15/24 09:20	11/15/24 13:38	EPA 200.7	
7439-92-1	Lead	1.57	U	1	1.57	6.00	ug/L	11/15/24 09:20	11/15/24 13:38	EPA 200.7	

Color Before: Colorless Clarity Before: Clear Texture:

Color After: Colorless Clarity After: Clear Artifacts:

Comments: Metals Group 10

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	P4853	OrderDate:	11/14/2024 11:45:00 AM
Client:	Tully Environmental, Inc	Project:	Transfer Station-SPDES
Contact:	Dean Devoe	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4853-01	001-WILLETS-PT-BLV D(NOV)	Water	Metals Group 10	200.7	11/13/24	11/15/24	11/15/24	11/14/24
P4853-02	002-35TH-AVE(NOV)	Water	Metals Group 10	200.7	11/13/24	11/15/24	11/15/24	11/14/24



SAMPLE DATA

Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	11/13/24 15:00
Project:	Transfer Station-SPDES	Date Received:	11/14/24
Client Sample ID:	001-WILLETS-PT-BLVD(NOV)	SDG No.:	P4853
Lab Sample ID:	P4853-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	0.40		1	0.045	0.10	mg/L	11/19/24 11:00	11/19/24 15:19	SM 4500-NH3 B plus G-11
BOD5	51.2		1	0.17	2.00	mg/L		11/14/24 16:30	SM 5210 B-16
Oil and Grease	15.7		1	0.40	5.00	mg/L		11/21/24 09:37	1664A
TSS	152		1	1.00	4.00	mg/L		11/15/24 09:00	SM 2540 D-15

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	11/13/24 15:00
Project:	Transfer Station-SPDES	Date Received:	11/14/24
Client Sample ID:	002-35TH-AVE(NOV)	SDG No.:	P4853
Lab Sample ID:	P4853-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	0.045	U	1	0.045	0.10	mg/L	11/19/24 11:00	11/19/24 15:19	SM 4500-NH3 B plus G-11
BOD5	44.2		1	0.17	2.00	mg/L		11/14/24 16:30	SM 5210 B-16
Oil and Grease	19.2		1	0.40	5.00	mg/L		11/21/24 09:37	1664A
TSS	133		1	1.00	4.00	mg/L		11/15/24 09:00	SM 2540 D-15

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	P4853	OrderDate:	11/14/2024 11:45:00 AM
Client:	Tully Environmental, Inc	Project:	Transfer Station-SPDES
Contact:	Dean Devoe	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4853-01	001-WILLETS-PT-BLV D(NOV)	WATER			11/13/24 15:00	11/19/24		11/14/24
			Ammonia	SM4500-NH3			11/19/24 15:19	
			BOD5	SM5210 B			11/14/24 16:30	
			Oil and Grease	1664A			11/21/24 09:37	
			TSS	SM2540 D			11/15/24 09:00	
P4853-02	002-35TH-AVE(NOV)	WATER			11/13/24 15:00	11/19/24		11/14/24
			Ammonia	SM4500-NH3			11/19/24 15:19	
			BOD5	SM5210 B			11/14/24 16:30	
			Oil and Grease	1664A			11/21/24 09:37	
			TSS	SM2540 D			11/15/24 09:00	



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Tully Environmental
 ADDRESS: 57 Seaview Blvd
 CITY: Pt Washington STATE: NY ZIP: 11050
 ATTENTION: D Devoe
 PHONE: 718 446 7000 FAX: 718 458 5199

CLIENT PROJECT INFORMATION

PROJECT NAME: Transfer Station SPAS
 PROJECT NO.: 242113 LOCATION:
 PROJECT MANAGER:
 e-mail:
 PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: Same PO#:
 ADDRESS:
 CITY STATE: ZIP:
 ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
 HARDCOPY (DATA PACKAGE): _____ DAYS*
 EDD: _____ DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
 + Raw Data ☐ Other _____
☐ EDD FORMAT _____

1 Cu Fe Pb
 2 BOD5
 3 TSS
 4 Hg
 5 1631LL
 6 BTEX
 7 D&G Ammonia
 8
 9

PRESERVATIVES

COMMENTS

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	001 Willets Pt Blvd (Nov)	W		X	11/13	3pm		X	X	X	X	X	X				
2.	002 35th Ave (Nov)	W		X	11/13	3pm		X	X	X	X	X	X				
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
1.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

QUISHED BY SAMPLER: D Devoe	DATE/TIME: 11/13/24	RECEIVED BY: 1. [Signature]	Conditions of bottles or coolers at receipt: <input type="checkbox"/> COMPLIANT <input type="checkbox"/> NON COMPLIANT <input type="checkbox"/> COOLER TEMP 4.0 °C
ISHED BY SAMPLER:	DATE/TIME: 1049	RECEIVED BY: 2. [Signature]	Comments: 7A Can #1
IED BY SAMPLER:	DATE/TIME: 11-14-24	RECEIVED BY: 3. [Signature]	

Page ____ of ____

CLIENT: ☐ Hand Delivered ☐ Other _____
 CHEMTECH: ☐ Picked Up ☐ Field Sampling

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : P4853	TULL01	Order Date : 11/14/2024 11:45:00 AM	Project Mgr :
Client Name : Tully Environmental, Inc		Project Name : Transfer Station-SPDES	Report Type : Results Only
Client Contact : Dean Devoe		Receive DateTime : 11/14/2024 10:49:00 AM	EDD Type : EXCEL NOCLEANUP
Invoice Name : Tully Environmental, Inc		Purchase Order :	Hard Copy Date :
Invoice Contact : Dean Devoe			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
P4853-01	001-WILLETS-PT-BLVD(NOV)	Water	11/13/2024	15:00	VOC-BTEX		624.1	5 Bus. Days	
P4853-02	002-35TH-AVE(NOV)	Water	11/13/2024	15:00	VOC-BTEX		624.1	5 Bus. Days	

Relinquished By :

Date / Time : 11-14-24 12:33

Received By :

Date / Time :

11-14-24 12:33

Storage Area : VOA Refridgerator Room