

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133471

Review By	kareem	Review On	11/19/2024 9:18:17 AM
Supervise By	jaswal	Supervise On	11/19/2024 10:27:58 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/15/24 10:42		Kareem	OK
2	S1	S1	CAL2	11/15/24 10:47		Kareem	OK
3	S2	S2	CAL3	11/15/24 10:51		Kareem	OK
4	S3	S3	CAL4	11/15/24 10:55		Kareem	OK
5	S4	S4	CAL5	11/15/24 10:59		Kareem	OK
6	S5	S5	CAL6	11/15/24 11:04		Kareem	OK
7	ICV01	ICV01	ICV	11/15/24 11:08	ICV fail for Ti,Zn (200.7)	Kareem	OK
8	LLICV01	LLICV01	LLICV	11/15/24 11:55		Kareem	OK
9	ICB01	ICB01	ICB	11/15/24 12:01		Kareem	OK
10	CRI01	CRI01	CRDL	11/15/24 12:05		Kareem	OK
11	ICSA01	ICSA01	ICSA	11/15/24 12:09		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	11/15/24 12:17		Kareem	OK
13	ICSADL	ICSADL	ICSA	11/15/24 12:22		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	11/15/24 12:26		Kareem	OK
15	CCV01	CCV01	CCV	11/15/24 12:30		Kareem	OK
16	CCB01	CCB01	CCB	11/15/24 12:38		Kareem	OK
17	PB164645BL	PB164645BL	MB	11/15/24 12:42	LCS fail for Na	Kareem	Not Ok
18	PB164645BS	PB164645BS	LCS	11/15/24 12:47	LCS fail for Na	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133471

Review By	kareem	Review On	11/19/2024 9:18:17 AM
Supervise By	jaswal	Supervise On	11/19/2024 10:27:58 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

19	P4834-03	SLP-1-STONE	SAM	11/15/24 12:51		Kareem	OK
20	P4835-03	345-2-STONE	SAM	11/15/24 12:55		Kareem	OK
21	P4836-03	T3-STONE	SAM	11/15/24 12:59		Kareem	OK
22	P4838-03	T1-STONE	SAM	11/15/24 13:03		Kareem	OK
23	P4837-01	3TRK-WATER	SAM	11/15/24 13:08		Kareem	OK
24	P4837-03	3TRK-STONE	SAM	11/15/24 13:12		Kareem	OK
25	P4853-01	001-WILLETS-PT-BL	SAM	11/15/24 13:33		Kareem	OK
26	P4853-02	002-35TH-AVE(NOV)	SAM	11/15/24 13:38		Kareem	OK
27	CCV02	CCV02	CCV	11/15/24 13:42		Kareem	OK
28	CCB02	CCB02	CCB	11/15/24 13:47		Kareem	OK
29	P4856-01	001-WILLETS-PT-BL	SAM	11/15/24 13:52		Kareem	OK
30	P4856-01DUP	001-WILLETS-PT-BL	DUP	11/15/24 13:56		Kareem	OK
31	P4856-01L	001-WILLETS-PT-BL	SD	11/15/24 14:00		Kareem	OK
32	P4856-01MS	001-WILLETS-PT-BL	MS	11/15/24 14:05	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
33	P4856-01MSD	001-WILLETS-PT-BL	MSD	11/15/24 14:09	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
34	P4856-01A	001-WILLETS-PT-BL	PS	11/15/24 14:13	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
35	P4856-02	002-35TH-AVE(OCT)	SAM	11/15/24 14:17		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133471

Review By	kareem	Review On	11/19/2024 9:18:17 AM
Supervise By	jaswal	Supervise On	11/19/2024 10:27:58 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

36	PB165002BL	PB165002BL	MB	11/15/24 14:29		Kareem	OK
37	CCV03	CCV03	CCV	11/15/24 14:40		Kareem	OK
38	CCB03	CCB03	CCB	11/15/24 14:44		Kareem	OK
39	P4837-01L	3TRK-WATERL	SD	11/15/24 14:49		Kareem	OK
40	P4837-01MS	3TRK-WATERMS	MS	11/15/24 14:53	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
41	P4837-01MSD	3TRK-WATERMSD	MSD	11/15/24 14:57	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
42	P4837-01A	3TRK-WATERA	PS	11/15/24 15:01	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
43	P4838-01	T1-WATER	SAM	11/15/24 15:05		Kareem	OK
44	PB164972BL	PB164972BL	MB	11/15/24 15:10		Kareem	OK
45	PB164972BS	PB164972BS	LCS	11/15/24 15:14	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
46	P4843-01	SW-WTS-01	SAM	11/15/24 15:18		Kareem	OK
47	P4824-01	T-20	SAM	11/15/24 15:23		Kareem	OK
48	P4825-01	STOCK-PILE	SAM	11/15/24 15:27		Kareem	OK
49	CCV004	CCV004	CCV	11/15/24 15:31		Kareem	OK
50	CCB004	CCB004	CCB	11/15/24 15:35		Kareem	OK
51	P4833-01	MH-731	SAM	11/15/24 15:40		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133471

Review By	kareem	Review On	11/19/2024 9:18:17 AM
Supervise By	jaswal	Supervise On	11/19/2024 10:27:58 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

52	P4833-05	TP-2	SAM	11/15/24 15:44		Kareem	OK
53	P4833-05DUP	TP-2DUP	DUP	11/15/24 15:48		Kareem	OK
54	P4833-05L	TP-2L	SD	11/15/24 15:52		Kareem	OK
55	P4833-05MS	TP-2MS	MS	11/15/24 15:57	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
56	P4833-05MSD	TP-2MSD	MSD	11/15/24 16:00	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
57	P4833-05A	TP-2A	PS	11/15/24 16:04	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
58	PB164970BL	PB164970BL	MB	11/15/24 16:08		Kareem	OK
59	PB164970BS	PB164970BS	LCS	11/15/24 16:13	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
60	CCV05	CCV05	CCV	11/15/24 16:17		Kareem	OK
61	CCB05	CCB05	CCB	11/15/24 16:21		Kareem	OK
62	P4821-04	BP-F-24	SAM	11/15/24 16:25		Kareem	OK
63	P4821-08	BP-F-2	SAM	11/15/24 16:30		Kareem	OK
64	P4823-04	MH-4	SAM	11/15/24 16:34		Kareem	OK
65	P4833-04	MH-731	SAM	11/15/24 16:38		Kareem	OK
66	P4833-08	TP-2	SAM	11/15/24 16:43		Kareem	OK
67	P4833-08DUP	TP-2DUP	DUP	11/15/24 16:47		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133471

Review By	kareem	Review On	11/19/2024 9:18:17 AM
Supervise By	jaswal	Supervise On	11/19/2024 10:27:58 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

68	P4833-08L	TP-2L	SD	11/15/24 16:52		Kareem	OK
69	P4833-08MS	TP-2MS	MS	11/15/24 16:56	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
70	P4833-08MSD	TP-2MSD	MSD	11/15/24 17:00	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
71	P4833-08A	TP-2A	PS	11/15/24 17:04	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
72	CCV06	CCV06	CCV	11/15/24 17:09		Kareem	OK
73	CCB06	CCB06	CCB	11/15/24 17:13		Kareem	OK
74	PB165002BS	PB165002BS	LCS	11/15/24 17:28	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
75	P4837-01DUP	3TRK-WATERDUP	DUP	11/15/24 17:32		Kareem	OK
76	PB164944TB	PB164944TB	MB	11/15/24 17:36	Cr fail	Kareem	Not Ok
77	PB164973BL	PB164973BL	MB	11/15/24 17:41		Kareem	OK
78	PB164973BS	PB164973BS	LCS	11/15/24 17:45	0.1 of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
79	LR1	LR1	HIGH STD	11/15/24 17:56		Kareem	OK
80	LR2	LR2	HIGH STD	11/15/24 18:01		Kareem	OK
81	CCV07	CCV07	CCV	11/15/24 18:08		Kareem	OK
82	CCB07	CCB07	CCB	11/15/24 18:12		Kareem	OK
83	P4813-04	419	SAM	11/15/24 18:16		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133471

Review By	kareem	Review On	11/19/2024 9:18:17 AM
Supervise By	jaswal	Supervise On	11/19/2024 10:27:58 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

84	CCV08	CCV08	CCV	11/15/24 18:21		Kareem	OK
85	CCB08	CCB08	CCB	11/15/24 18:25		Kareem	OK