

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: P4853 SAS No.: P4853 SDG No.: P4853

Instrument ID: MSVOA_N Calibration Date/Time: 11/15/2024 13:49

Lab File ID: VN084867.D Init. Calib. Date(s): 10/31/2024 10/31/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:08 18:13

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.494	1.459	0.5	-2.34	
Toluene	1.709	1.664	0.4	-2.63	
Ethyl Benzene	1.796	1.629	0.1	-9.3	
m/p-Xylenes	0.687	0.652	0.3	-5.09	
o-Xylene	0.663	0.629	0.3	-5.13	
1,2-Dichloroethane-d4	2.412	2.422	0.01	0.41	
Toluene-d8	1.420	1.438	0.01	1.27	
4-Bromofluorobenzene	0.517	0.524	0.2	1.35	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.