

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: P4853 SAS No.: P4853 SDG No.: P4853

Instrument ID: MSVOA_N Calibration Date/Time: 11/21/2024 10:09

Lab File ID: VN084987.D Init. Calib. Date(s): 10/31/2024 10/31/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:08 18:13

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.494	1.418	0.5	-5.09	
Toluene	1.709	1.649	0.4	-3.51	
Ethyl Benzene	1.796	1.629	0.1	-9.3	
m/p-Xylenes	0.687	0.661	0.3	-3.79	
o-Xylene	0.663	0.630	0.3	-4.98	
1,2-Dichloroethane-d4	2.412	2.441	0.01	1.2	
Toluene-d8	1.420	1.421	0.01	0.07	
4-Bromofluorobenzene	0.517	0.540	0.2	4.45	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.