

Surrogate Summary

SDG No.: **P4853**

Client: **Tully Environmental, Inc**

Analytical Method: **SW624.1**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4853-01	001-WILLETS-PT-BLVD(NOV)	1,2-Dichloroethane-d4	30	31.3	104	91	110
		Toluene-d8	30	28.5	95	91	112
		4-Bromofluorobenzene	30	25.4	85	63	112
P4853-01DL	001-WILLETS-PT-BLVD(NOV)DL	1,2-Dichloroethane-d4	30	30.8	103	91	110
		Toluene-d8	30	28.4	95	91	112
		4-Bromofluorobenzene	30	25.8	86	63	112
P4853-02	002-35TH-AVE(NOV)	1,2-Dichloroethane-d4	30	30.8	103	91	110
		Toluene-d8	30	28.2	94	91	112
		4-Bromofluorobenzene	30	26.4	88	63	112
P4853-02DL	002-35TH-AVE(NOV)DL	1,2-Dichloroethane-d4	30	30.1	100	91	110
		Toluene-d8	30	28.0	93	91	112
		4-Bromofluorobenzene	30	25.2	84	63	112
VN1115WBL01	VN1115WBL01	1,2-Dichloroethane-d4	30	29.4	98	91	110
		Toluene-d8	30	28.4	95	91	112
		4-Bromofluorobenzene	30	23.9	80	63	112
VN1115WBS01	VN1115WBS01	1,2-Dichloroethane-d4	30	30.7	102	91	110
		Toluene-d8	30	29.7	99	91	112
		4-Bromofluorobenzene	30	30.2	101	63	112
VN1115WBSD0	VN1115WBSD01	1,2-Dichloroethane-d4	30	29.9	100	91	110
		Toluene-d8	30	29.6	99	91	112
		4-Bromofluorobenzene	30	31.0	103	63	112
VN1121WBL01	VN1121WBL01	1,2-Dichloroethane-d4	30	30.9	103	91	110
		Toluene-d8	30	28.0	93	91	112
		4-Bromofluorobenzene	30	24.4	81	63	112
VN1121WBS01	VN1121WBS01	1,2-Dichloroethane-d4	30	29.8	99	91	110
		Toluene-d8	30	29.8	99	91	112
		4-Bromofluorobenzene	30	29.9	100	63	112