

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4853

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VN084988.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN1121WBS01	Benzene	20	19.2	ug/L	96			65	135	
	Toluene	20	19.3	ug/L	97			70	130	
	Ethyl Benzene	20	18.7	ug/L	94			60	140	
	m/p-Xylenes	40	39.3	ug/L	98			87	111	
	o-Xylene	20	18.8	ug/L	94			87	111	