

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
Lab Code: CHEM Case No.: P4853 SAS No.: P4853 SDG NO.: P4853
Lab File ID: VN084867.D Date Analyzed: 11/15/2024
Instrument ID: MSVOA_N Time Analyzed: 13:49
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	29205	7.81	143711	9.10	129619	11.87
UPPER LIMIT	58410	8.312	287422	9.6	259238	12.365
LOWER LIMIT	14602.5	7.312	71855.5	8.6	64809.5	11.365
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (NOV)	22235	7.81	116438	9.10	107920	11.87
002-35TH-AVE (NOV)	23575	7.81	122982	9.10	114966	11.87
VN1115WBL01	20903	7.81	102633	9.10	89042	11.87
VN1115WBS01	28727	7.81	152762	9.10	143052	11.87
VN1115WBSD01	27777	7.81	143562	9.10	132133	11.87

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.