



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **Tetra Tech**

**Project Name: Ocean County Landfill 2024**

**Project # N/A**

**Chemtech Project # P4860**

**Test Name: SVOC-TCL BNA -20**

### **A. Number of Samples and Date of Receipt:**

10 Solid samples were received on 11/14/2024.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Cyanide, EPH, Hexavalent Chromium, Mercury, Metals ICP-TAL, PCB, Pesticide-TCL, pH, SVOC-TCL BNA -20, TCL+30/TAL and VOC-TCL VOA-10. This data package contains results for SVOC-TCL BNA -20.

### **C. Analytical Techniques:**

The samples were analyzed on instrument BNA\_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The analysis of SVOC-TCL BNA -20 was based on method 8270E and extraction was done based on method 3541.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {P4860-09MS} with File ID: BF140639.D recoveries met the requirements for all compounds except for 3,3-Dichlorobenzidine[65%], 3-Nitroaniline[55%], 4-Chloroaniline[40%] and Benzaldehyde[12%] these compounds did not meet the NJDKQP criteria but met the in-house criteria .

The MSD {P4860-09MSD} with File ID: BF140640.D recoveries met the acceptable requirements except for 3-Nitroaniline[57%], 4-Chloroaniline[40%] and Benzaldehyde[12%] these compounds did not meet the NJDKQP criteria but met the in-house criteria .

The RPD met criteria .

The Blank Spike for {PB165029BS} with File ID: BF140445.D met requirements for all samples except for 3,3-Dichlorobenzidine[55%], 3-Nitroaniline[54%], 4-Chloroaniline[48%], Benzaldehyde[11%] these compounds did not meet the NJDKQP criteria but met the in-house criteria while Hexachlorocyclopentadiene[173%] this compound did not meet the NJDKQP criteria and in-house criteria, is failing high but no positive hit in associate sample therefore no corrective action taken.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 20% in the Initial Calibration (8270-BF112124.M) for Hexachlorocyclopentadiene, 2,4-Dinitrophenol, this compound is passing on Linear Regression.

The Continuous Calibration met the requirements .  
The Tuning criteria met requirements.

**E. Additional Comments:**

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_