



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Tetra Tech

SDG No.: P4860

Project: Ocean County Landfill 2024

RunNo.: LB133478

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	11/15/2024
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	11/15/2024
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	11/15/2024
Sample ID: CCV3 pH	pH	2.01	2.00	101	90-110	11/15/2024

Initial and Continuing Calibration Verification

Client: Tetra Tech

SDG No.: P4860

Project: Ocean County Landfill 2024

RunNo.: LB133487

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
Cyanide		mg/L	0.097	0.099	98	90-110	11/15/2024
Sample ID:	CCV1						
Cyanide		mg/L	0.25	0.25	100	90-110	11/15/2024
Sample ID:	CCV2						
Cyanide		mg/L	0.25	0.25	100	90-110	11/15/2024
Sample ID:	CCV3						
Cyanide		mg/L	0.25	0.25	100	90-110	11/15/2024
Sample ID:	CCV4						
Cyanide		mg/L	0.25	0.25	100	90-110	11/15/2024

Initial and Continuing Calibration Verification

Client: Tetra Tech

SDG No.: P4860

Project: Ocean County Landfill 2024

RunNo.: LB133497

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	11/18/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	11/18/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	11/18/2024
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	11/18/2024
Sample ID: CCV4 Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	11/18/2024