

Laboratory Control Sample Summary

Client: Tetra Tech

SDG No.: P4860

Project: Ocean County Landfill 2024

Run No.: LB133487

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165016BS							
Cyanide	mg/Kg	5	4.80		96	1	85-115	11/15/2024

Laboratory Control Sample Summary

Client: Tetra Tech

SDG No.: P4860

Project: Ocean County Landfill 2024

Run No.: LB133497

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165024BS							
Hexavalent Chromium	mg/Kg	20	20.1		101	1	84-110	11/18/2024