

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133502

Review By	kareem	Review On	11/19/2024 9:23:02 AM
Supervise By	jaswal	Supervise On	11/19/2024 10:28:07 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/18/24 12:31		Kareem	OK
2	S1	S1	CAL2	11/18/24 12:36		Kareem	OK
3	S2	S2	CAL3	11/18/24 12:40		Kareem	OK
4	S3	S3	CAL4	11/18/24 12:44		Kareem	OK
5	S4	S4	CAL5	11/18/24 12:49		Kareem	OK
6	S5	S5	CAL6	11/18/24 12:53		Kareem	OK
7	ICV01	ICV01	ICV	11/18/24 12:57		Kareem	OK
8	LLICV01	LLICV01	LLICV	11/18/24 13:11		Kareem	OK
9	ICB01	ICB01	ICB	11/18/24 13:15		Kareem	OK
10	CRI01	CRI01	CRDL	11/18/24 13:20		Kareem	OK
11	ICSA01	ICSA01	ICSA	11/18/24 13:30		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	11/18/24 13:42		Kareem	OK
13	CCV01	CCV01	CCV	11/18/24 13:52		Kareem	OK
14	CCB01	CCB01	CCB	11/18/24 14:02		Kareem	OK
15	P4857-01	ETGI-328	SAM	11/18/24 14:06		Kareem	OK
16	P4857-03	VNJ-226	SAM	11/18/24 14:10		Kareem	OK
17	P4859-01	NB-08-111424	SAM	11/18/24 14:14		Kareem	OK
18	P4867-02	TT-070-IDWGW-2024	SAM	11/18/24 14:18		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133502

Review By	kareem	Review On	11/19/2024 9:23:02 AM
Supervise By	jaswal	Supervise On	11/19/2024 10:28:07 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

19	P4867-03	TT-071-IDWGW-2024	SAM	11/18/24 14:23		Kareem	OK
20	P4867-04	TT-072-IDWGW-2024	SAM	11/18/24 14:27		Kareem	OK
21	CCV02	CCV02	CCV	11/18/24 14:31		Kareem	OK
22	CCB02	CCB02	CCB	11/18/24 14:35		Kareem	OK
23	PB165011BL	PB165011BL	MB	11/18/24 14:40	Not Use	Kareem	Not Ok
24	PB165011BS	PB165011BS	LCS	11/18/24 14:45	Not Use	Kareem	Not Ok
25	PB165003BL	PB165003BL	MB	11/18/24 14:49		Kareem	OK
26	PB165003BS	PB165003BS	LCS	11/18/24 14:53	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
27	P4848-01	TP-1	SAM	11/18/24 15:10	Cr high	Kareem	Dilution
28	P4849-01	RR-1	SAM	11/18/24 15:14		Kareem	OK
29	P4850-01	OR-03-11142024	SAM	11/18/24 15:18		Kareem	OK
30	P4851-01	OK-02-11142024	SAM	11/18/24 15:22		Kareem	OK
31	P4852-01	COMP-1	SAM	11/18/24 15:26		Kareem	OK
32	P4852-02	COMP-2	SAM	11/18/24 15:31		Kareem	OK
33	CCV03	CCV03	CCV	11/18/24 15:35		Kareem	OK
34	CCB03	CCB03	CCB	11/18/24 15:39		Kareem	OK
35	P4852-03	COMP-3	SAM	11/18/24 15:43		Kareem	OK
36	P4859-01DUP	NB-08-111424DUP	DUP	11/18/24 15:48		Kareem	OK
37	P4859-01L	NB-08-111424L	SD	11/18/24 15:52		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133502

Review By	kareem	Review On	11/19/2024 9:23:02 AM
Supervise By	jaswal	Supervise On	11/19/2024 10:28:07 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

38	P4859-01MS	NB-08-111424MS	MS	11/18/24 15:56	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
39	P4859-01MSD	NB-08-111424MSD	MSD	11/18/24 16:00	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
40	P4859-01A	NB-08-111424A	PS	11/18/24 16:04	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
41	P4868-01	RW5-SP100-2024111	SAM	11/18/24 16:08		Kareem	OK
42	P4868-03	RW5-SP303-2024111	SAM	11/18/24 16:12		Kareem	OK
43	P4868-03DUP	RW5-SP303-2024111	DUP	11/18/24 16:16		Kareem	OK
44	P4868-03MSD	RW5-SP303-2024111	MSD	11/18/24 16:21	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
45	CCV04	CCV04	CCV	11/18/24 16:25		Kareem	OK
46	CCB04	CCB04	CCB	11/18/24 16:29		Kareem	OK
47	P4848-01DL	TP-1DL	SAM	11/18/24 16:33	10x for Cr	Kareem	Confirms
48	P4868-03A	RW5-SP303-2024111	PS	11/18/24 16:37	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
49	P4868-03L	RW5-SP303-2024111	SD	11/18/24 16:41		Kareem	OK
50	P4868-03MS	RW5-SP303-2024111	MS	11/18/24 16:46	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
51	LR1	LR1	HIGH STD	11/18/24 16:51		Kareem	OK
52	LR2	LR2	HIGH STD	11/18/24 16:56		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133502

Review By	kareem	Review On	11/19/2024 9:23:02 AM
Supervise By	jaswal	Supervise On	11/19/2024 10:28:07 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

53	CCV05	CCV05	CCV	11/18/24 17:01		Kareem	OK
54	CCB05	CCB05	CCB	11/18/24 17:06		Kareem	OK
55	P4861-01	WC-11-A-202411	SAM	11/18/24 17:27		Kareem	OK
56	P4861-01DUP	WC-11-A-202411DUP	DUP	11/18/24 17:31		Kareem	OK
57	P4861-01L	WC-11-A-202411L	SD	11/18/24 17:35		Kareem	OK
58	P4861-01MS	WC-11-A-202411MS	MS	11/18/24 17:40	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
59	P4861-01MSD	WC-11-A-202411MSD	MSD	11/18/24 17:44	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
60	P4861-01A	WC-11-A-202411A	PS	11/18/24 17:48	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
61	P4871-01	WC-10-A-202411	SAM	11/18/24 17:51		Kareem	OK
62	P4848-02	TP-1	SAM	11/18/24 17:56		Kareem	OK
63	P4869-01	EO-02-11152024	SAM	11/18/24 18:00		Kareem	OK
64	CCV06	CCV06	CCV	11/18/24 18:04		Kareem	OK
65	CCB06	CCB06	CCB	11/18/24 18:08		Kareem	OK
66	P4887-01	MH-739	SAM	11/18/24 18:13		Kareem	OK
67	P4887-05	MH-760	SAM	11/18/24 18:17		Kareem	OK
68	P4889-01	VNJ-228	SAM	11/18/24 18:21		Kareem	OK
69	P4889-03	#72-11978	SAM	11/18/24 18:25		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133502

Review By	kareem	Review On	11/19/2024 9:23:02 AM
Supervise By	jaswal	Supervise On	11/19/2024 10:28:07 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

70	P4889-05	#72-11930	SAM	11/18/24 18:29		Kareem	OK
71	P4889-05DUP	#72-11930DUP	DUP	11/18/24 18:33		Kareem	OK
72	P4889-05L	#72-11930L	SD	11/18/24 18:37		Kareem	OK
73	P4889-05MS	#72-11930MS	MS	11/18/24 18:42	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
74	P4889-05MSD	#72-11930MSD	MSD	11/18/24 18:46	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
75	P4889-05A	#72-11930A	PS	11/18/24 18:50	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
76	CCV07	CCV07	CCV	11/18/24 18:54		Kareem	OK
77	CCB07	CCB07	CCB	11/18/24 18:58		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133524

Review By	kareem	Review On	11/20/2024 11:25:14 AM
Supervise By	Jaswal	Supervise On	11/20/2024 11:38:44 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/19/24 10:55		Kareem	OK
2	S1	S1	CAL2	11/19/24 11:00		Kareem	OK
3	S2	S2	CAL3	11/19/24 11:04		Kareem	OK
4	S3	S3	CAL4	11/19/24 11:08		Kareem	OK
5	S4	S4	CAL5	11/19/24 11:13		Kareem	OK
6	S5	S5	CAL6	11/19/24 11:17		Kareem	OK
7	ICV01	ICV01	ICV	11/19/24 11:25		Kareem	OK
8	LLICV01	LLICV01	LLICV	11/19/24 11:34		Kareem	OK
9	ICB01	ICB01	ICB	11/19/24 11:38		Kareem	OK
10	CRI01	CRI01	CRDL	11/19/24 11:43		Kareem	OK
11	ICSA01	ICSA01	ICSA	11/19/24 11:47		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	11/19/24 11:51		Kareem	OK
13	CCV01	CCV01	CCV	11/19/24 12:43		Kareem	OK
14	CCB01	CCB01	CCB	11/19/24 12:48		Kareem	OK
15	PB165067BL	PB165067BL	MB	11/19/24 12:52		Kareem	OK
16	PB165067BS	PB165067BS	LCS	11/19/24 12:56	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
17	PB164759BL	PB164759BL	MB	11/19/24 13:00		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133524

Review By	kareem	Review On	11/20/2024 11:25:14 AM
Supervise By	Jaswal	Supervise On	11/20/2024 11:38:44 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

18	PB164759BS	PB164759BS	LCS	11/19/24 13:05	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
19	PB164645BL	PB164645BL	MB	11/19/24 13:09		Kareem	OK
20	PB164645BS	PB164645BS	LCS	11/19/24 13:13	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
21	PB164944TB	PB164944TB	MB	11/19/24 13:17		Kareem	OK
22	P4892-01	WB-310-TOP	SAM	11/19/24 13:22	Cr high	Kareem	Dilution
23	P4892-02	WB-310-BOT	SAM	11/19/24 13:26		Kareem	OK
24	P4893-01	MH-763	SAM	11/19/24 13:30		Kareem	OK
25	CCV02	CCV02	CCV	11/19/24 13:34		Kareem	OK
26	CCB02	CCB02	CCB	11/19/24 13:38		Kareem	OK
27	P4893-05	MH-762	SAM	11/19/24 13:42		Kareem	OK
28	P4908-01	SP-1	SAM	11/19/24 13:46		Kareem	OK
29	P4908-03	SP-2	SAM	11/19/24 13:51		Kareem	OK
30	P4909-01	BU-02-111824	SAM	11/19/24 13:55		Kareem	OK
31	P4909-01DUP	BU-02-111824DUP	DUP	11/19/24 13:59		Kareem	OK
32	P4909-01L	BU-02-111824L	SD	11/19/24 14:03		Kareem	OK
33	P4909-01MS	BU-02-111824MS	MS	11/19/24 14:08	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
34	P4909-01MSD	BU-02-111824MSD	MSD	11/19/24 14:12	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133524

Review By	kareem	Review On	11/20/2024 11:25:14 AM
Supervise By	Jaswal	Supervise On	11/20/2024 11:38:44 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

35	P4909-01A	BU-02-111824A	PS	11/19/24 14:16	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
36	P4890-01	#3711	SAM	11/19/24 14:20		Kareem	OK
37	CCV03	CCV03	CCV	11/19/24 14:24		Kareem	OK
38	CCB03	CCB03	CCB	11/19/24 14:28		Kareem	OK
39	P4890-02	D3617	SAM	11/19/24 14:37		Kareem	OK
40	P4890-03	D3690	SAM	11/19/24 14:41		Kareem	OK
41	P4849-02	RR-1	SAM	11/19/24 14:56		Kareem	OK
42	PB165019TB	PB165019TB	MB	11/19/24 15:00		Kareem	OK
43	P4870-13	TP-1	SAM	11/19/24 15:05		Kareem	OK
44	P4870-14	MH-735	SAM	11/19/24 15:09		Kareem	OK
45	P4870-15	MH-736	SAM	11/19/24 15:13		Kareem	OK
46	P4870-16	TP-15	SAM	11/19/24 15:18		Kareem	OK
47	P4887-02	MH-739	SAM	11/19/24 15:22		Kareem	OK
48	P4887-06	MH-760	SAM	11/19/24 15:27		Kareem	OK
49	CCV04	CCV04	CCV	11/19/24 15:31		Kareem	OK
50	CCB04	CCB04	CCB	11/19/24 15:36		Kareem	OK
51	PB164986TB	PB164986TB	MB	11/19/24 15:40	fail for Cr	Kareem	Not Ok
52	PB165049BL	PB165049BL	MB	11/19/24 15:44		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133524

Review By	kareem	Review On	11/20/2024 11:25:14 AM
Supervise By	Jaswal	Supervise On	11/20/2024 11:38:44 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

53	PB165049BS	PB165049BS	LCS	11/19/24 15:49	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
54	P4892-01DL	WB-310-TOPDL	SAM	11/19/24 15:53	5x for Cr	Kareem	Confirms
55	P4341-02	10324-0	SAM	11/19/24 15:57		Kareem	OK
56	P4860-01	DUP-01	SAM	11/19/24 16:01		Kareem	OK
57	P4860-02	PH2-BOT-001	SAM	11/19/24 16:06		Kareem	OK
58	P4860-03	PH2-BOT-002	SAM	11/19/24 16:10		Kareem	OK
59	P4860-04	PH2-BOT-003	SAM	11/19/24 16:14		Kareem	OK
60	CCV05	CCV05	CCV	11/19/24 16:18		Kareem	OK
61	CCB05	CCB05	CCB	11/19/24 16:22		Kareem	OK
62	P4860-05	PH2-BOT-004	SAM	11/19/24 16:27		Kareem	OK
63	P4860-06	PH2-BOT-009	SAM	11/19/24 16:31		Kareem	OK
64	P4860-07	PH2-BOT-008	SAM	11/19/24 16:35		Kareem	OK
65	P4860-08	PH2-BOT-007	SAM	11/19/24 16:39		Kareem	OK
66	P4860-09	PH2-BOT-006	SAM	11/19/24 16:44		Kareem	OK
67	P4860-10	PH2-BOT-005	SAM	11/19/24 16:48		Kareem	OK
68	P4870-01	TP-1	SAM	11/19/24 16:52		Kareem	OK
69	P4870-04	MH-735	SAM	11/19/24 16:57		Kareem	OK
70	P4870-07	MH-736	SAM	11/19/24 17:01		Kareem	OK
71	P4870-10	TP-15	SAM	11/19/24 17:05		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133524

Review By	kareem	Review On	11/20/2024 11:25:14 AM
Supervise By	Jaswal	Supervise On	11/20/2024 11:38:44 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

72	CCV06	CCV06	CCV	11/19/24 17:19		Kareem	OK
73	CCB06	CCB06	CCB	11/19/24 17:24		Kareem	OK
74	PB165047BL	PB165047BL	MB	11/19/24 17:28		Kareem	OK
75	PB165047BS	PB165047BS	LCS	11/19/24 17:32		Kareem	OK
76	LR1	LR1	HIGH STD	11/19/24 17:51		Kareem	OK
77	LR2	LR2	HIGH STD	11/19/24 17:56		Kareem	OK
78	CCV07	CCV07	CCV	11/19/24 18:01		Kareem	OK
79	CCB07	CCB07	CCB	11/19/24 18:05		Kareem	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133581

Review By	jaswal	Review On	11/22/2024 9:51:42 PM
Supervise By	mohan	Supervise On	11/22/2024 9:53:32 PM

STD. NAME	STD REF.#
ICAL Standard	MP83330,MP83331,MP83332,MP83333,MP83334,MP83335
ICV Standard	MP83336
CCV Standard	MP83338
ICSA Standard	
CRI Standard	MP83340
LCS Standard	
Chk Standard	MP83337,MP83339,MP83341,MP83345

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/22/24 13:37		Mohan	OK
2	S0.2	S0.2	CAL2	11/22/24 13:40		Mohan	OK
3	S2.5	S2.5	CAL3	11/22/24 13:42		Mohan	OK
4	S5	S5	CAL4	11/22/24 13:44		Mohan	OK
5	S7.5	S7.5	CAL5	11/22/24 13:49		Mohan	OK
6	S10	S10	CAL6	11/22/24 13:55		Mohan	OK
7	ICV98	ICV98	ICV	11/22/24 13:57		Mohan	OK
8	ICB98	ICB98	ICB	11/22/24 14:00		Mohan	OK
9	CCV37	CCV37	CCV	11/22/24 14:02		Mohan	OK
10	CCB37	CCB37	CCB	11/22/24 14:04		Mohan	OK
11	CRA	CRA	CRDL	11/22/24 14:07		Mohan	OK
12	HighStd	HighStd	HIGH STD	11/22/24 14:09		Mohan	OK
13	ChkStd	ChkStd	SAM	11/22/24 14:11		Mohan	OK
14	PB165177BL	PB165177BL	MB	11/22/24 14:13		Mohan	OK
15	PB165177BS	PB165177BS	LCS	11/22/24 14:16		Mohan	OK
16	P4860-01	DUP-01	SAM	11/22/24 14:18		Mohan	OK
17	P4860-01DUP	DUP-01DUP	DUP	11/22/24 14:20		Mohan	OK
18	P4860-01MS	DUP-01MS	MS	11/22/24 14:23		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133581

Review By	jaswal	Review On	11/22/2024 9:51:42 PM
Supervise By	mohan	Supervise On	11/22/2024 9:53:32 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83330,MP83331,MP83332,MP83333,MP83334,MP83335		
ICV Standard	MP83336		
CCV Standard	MP83338		
ICSA Standard			
CRI Standard	MP83340		
LCS Standard			
Chk Standard	MP83337,MP83339,MP83341,MP83345		

19	P4860-01MSD	DUP-01MSD	MSD	11/22/24 14:25		Mohan	OK
20	P4860-02	PH2-BOT-001	SAM	11/22/24 14:27		Mohan	OK
21	CCV38	CCV38	CCV	11/22/24 14:30		Mohan	OK
22	CCB38	CCB38	CCB	11/22/24 14:32		Mohan	OK
23	P4860-03	PH2-BOT-002	SAM	11/22/24 14:34		Mohan	OK
24	P4860-04	PH2-BOT-003	SAM	11/22/24 14:36		Mohan	OK
25	P4860-05	PH2-BOT-004	SAM	11/22/24 14:39		Mohan	OK
26	P4860-06	PH2-BOT-009	SAM	11/22/24 14:41		Mohan	OK
27	P4860-07	PH2-BOT-008	SAM	11/22/24 14:43		Mohan	OK
28	P4860-08	PH2-BOT-007	SAM	11/22/24 14:46		Mohan	OK
29	P4860-09	PH2-BOT-006	SAM	11/22/24 14:48		Mohan	OK
30	P4860-10	PH2-BOT-005	SAM	11/22/24 14:50		Mohan	OK
31	PB165178BL	PB165178BL	MB	11/22/24 14:52		Mohan	OK
32	PB165178BS	PB165178BS	LCS	11/22/24 14:55		Mohan	OK
33	CCV39	CCV39	CCV	11/22/24 14:57		Mohan	OK
34	CCB39	CCB39	CCB	11/22/24 14:59		Mohan	OK
35	P4916-01	TP-1-WC	SAM	11/22/24 15:02		Mohan	OK
36	P4916-05	TP-2-WC	SAM	11/22/24 15:04		Mohan	OK
37	P4916-09	TP-3-WC	SAM	11/22/24 15:06		Mohan	OK
38	P4923-06	72-11991	SAM	11/22/24 15:08		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133581

Review By	jaswal	Review On	11/22/2024 9:51:42 PM
Supervise By	mohan	Supervise On	11/22/2024 9:53:32 PM

STD. NAME	STD REF.#
ICAL Standard	MP83330,MP83331,MP83332,MP83333,MP83334,MP83335
ICV Standard	MP83336
CCV Standard	MP83338
ICSA Standard	
CRI Standard	MP83340
LCS Standard	
Chk Standard	MP83337,MP83339,MP83341,MP83345

39	P4924-01	MH-4	SAM	11/22/24 15:11		Mohan	OK
40	P4925-01	MH-741	SAM	11/22/24 15:13		Mohan	OK
41	P4925-05	MH-758	SAM	11/22/24 15:15		Mohan	OK
42	P4929-01	ARS520	SAM	11/22/24 15:17		Mohan	OK
43	P4936-01	PL-01-11202024	SAM	11/22/24 15:20		Mohan	OK
44	P4938-01	MH-732	SAM	11/22/24 15:22		Mohan	OK
45	CCV40	CCV40	CCV	11/22/24 15:24		Mohan	OK
46	CCB40	CCB40	CCB	11/22/24 15:26		Mohan	OK
47	P4938-05	MH-734	SAM	11/22/24 15:29		Mohan	OK
48	P4938-05DUP	MH-734DUP	DUP	11/22/24 15:31		Mohan	OK
49	P4938-05MS	MH-734MS	MS	11/22/24 15:33		Mohan	OK
50	P4938-05MSD	MH-734MSD	MSD	11/22/24 15:36		Mohan	OK
51	P4860-01L	DUP-01L	SD	11/22/24 15:38		Mohan	OK
52	P4860-01A	DUP-01A	PS	11/22/24 15:40		Mohan	OK
53	P4938-05L	MH-734L	SD	11/22/24 15:43		Mohan	OK
54	P4938-05A	MH-734A	PS	11/22/24 15:45		Mohan	OK
55	CCV41	CCV41	CCV	11/22/24 15:47		Mohan	OK
56	CCB41	CCB41	CCB	11/22/24 15:49		Mohan	OK