



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Tetra Tech				SDG No.:	P4860				
Contract:	CORN02		Lab Code:	CHEM		Case No.:	P4860		SAS No.:	P4860
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB98	Mercury	0.20	+/-0.20	U	0.20	CV	11/22/2024	14:00	LB133581	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech

SDG No.: P4860

Contract: CORN02

Lab Code: CHEM

Case No.: P4860

SAS No.: P4860

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB37	Mercury	0.20	+/-0.20	U	0.20	CV	11/22/2024	14:04	LB133581
CCB38	Mercury	0.20	+/-0.20	U	0.20	CV	11/22/2024	14:32	LB133581
CCB39	Mercury	0.20	+/-0.20	U	0.20	CV	11/22/2024	14:59	LB133581
CCB40	Mercury	0.20	+/-0.20	U	0.20	CV	11/22/2024	15:26	LB133581
CCB41	Mercury	0.20	+/-0.20	U	0.20	CV	11/22/2024	15:49	LB133581

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech</u>		SDG No.: <u>P4860</u>							
Contract: <u>CORN02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4860</u>	SAS No.: <u>P4860</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/18/2024	13:15	LB133502
	Antimony	50.0	+/-50.0	U	50.0	P	11/18/2024	13:15	LB133502
	Arsenic	20.0	+/-20.0	U	20.0	P	11/18/2024	13:15	LB133502
	Barium	100	+/-100	U	100	P	11/18/2024	13:15	LB133502
	Beryllium	6.00	+/-6.00	U	6.00	P	11/18/2024	13:15	LB133502
	Cadmium	6.00	+/-6.00	U	6.00	P	11/18/2024	13:15	LB133502
	Calcium	2000	+/-2000	U	2000	P	11/18/2024	13:15	LB133502
	Chromium	10.0	+/-10.0	U	10.0	P	11/18/2024	13:15	LB133502
	Cobalt	30.0	+/-30.0	U	30.0	P	11/18/2024	13:15	LB133502
	Copper	20.0	+/-20.0	U	20.0	P	11/18/2024	13:15	LB133502
	Iron	100	+/-100	U	100	P	11/18/2024	13:15	LB133502
	Lead	12.0	+/-12.0	U	12.0	P	11/18/2024	13:15	LB133502
	Magnesium	2000	+/-2000	U	2000	P	11/18/2024	13:15	LB133502
	Manganese	20.0	+/-20.0	U	20.0	P	11/18/2024	13:15	LB133502
	Nickel	40.0	+/-40.0	U	40.0	P	11/18/2024	13:15	LB133502
	Potassium	2000	+/-2000	U	2000	P	11/18/2024	13:15	LB133502
	Selenium	20.0	+/-20.0	U	20.0	P	11/18/2024	13:15	LB133502
	Silver	10.0	+/-10.0	U	10.0	P	11/18/2024	13:15	LB133502
	Sodium	2000	+/-2000	U	2000	P	11/18/2024	13:15	LB133502
	Thallium	40.0	+/-40.0	U	40.0	P	11/18/2024	13:15	LB133502
	Vanadium	40.0	+/-40.0	U	40.0	P	11/18/2024	13:15	LB133502
	Zinc	40.0	+/-40.0	U	40.0	P	11/18/2024	13:15	LB133502

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech</u>		SDG No.: <u>P4860</u>							
Contract: <u>CORN02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4860</u>	SAS No.: <u>P4860</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/18/2024	14:02	LB133502
	Antimony	50.0	+/-50.0	U	50.0	P	11/18/2024	14:02	LB133502
	Arsenic	20.0	+/-20.0	U	20.0	P	11/18/2024	14:02	LB133502
	Barium	100	+/-100	U	100	P	11/18/2024	14:02	LB133502
	Beryllium	6.00	+/-6.00	U	6.00	P	11/18/2024	14:02	LB133502
	Cadmium	6.00	+/-6.00	U	6.00	P	11/18/2024	14:02	LB133502
	Calcium	2000	+/-2000	U	2000	P	11/18/2024	14:02	LB133502
	Chromium	10.0	+/-10.0	U	10.0	P	11/18/2024	14:02	LB133502
	Cobalt	30.0	+/-30.0	U	30.0	P	11/18/2024	14:02	LB133502
	Copper	20.0	+/-20.0	U	20.0	P	11/18/2024	14:02	LB133502
	Iron	100	+/-100	U	100	P	11/18/2024	14:02	LB133502
	Lead	12.0	+/-12.0	U	12.0	P	11/18/2024	14:02	LB133502
	Magnesium	2000	+/-2000	U	2000	P	11/18/2024	14:02	LB133502
	Manganese	20.0	+/-20.0	U	20.0	P	11/18/2024	14:02	LB133502
	Nickel	40.0	+/-40.0	U	40.0	P	11/18/2024	14:02	LB133502
	Potassium	2000	+/-2000	U	2000	P	11/18/2024	14:02	LB133502
	Selenium	20.0	+/-20.0	U	20.0	P	11/18/2024	14:02	LB133502
	Silver	10.0	+/-10.0	U	10.0	P	11/18/2024	14:02	LB133502
	Sodium	2000	+/-2000	U	2000	P	11/18/2024	14:02	LB133502
	Thallium	40.0	+/-40.0	U	40.0	P	11/18/2024	14:02	LB133502
	Vanadium	40.0	+/-40.0	U	40.0	P	11/18/2024	14:02	LB133502
	Zinc	40.0	+/-40.0	U	40.0	P	11/18/2024	14:02	LB133502
CCB02	Aluminum	100	+/-100	U	100	P	11/18/2024	14:35	LB133502
	Antimony	4.42	+/-50.0	J	50.0	P	11/18/2024	14:35	LB133502
	Arsenic	20.0	+/-20.0	U	20.0	P	11/18/2024	14:35	LB133502
	Barium	18.0	+/-100	J	100	P	11/18/2024	14:35	LB133502
	Beryllium	6.00	+/-6.00	U	6.00	P	11/18/2024	14:35	LB133502
	Cadmium	6.00	+/-6.00	U	6.00	P	11/18/2024	14:35	LB133502
	Calcium	2000	+/-2000	U	2000	P	11/18/2024	14:35	LB133502
	Chromium	10.0	+/-10.0	U	10.0	P	11/18/2024	14:35	LB133502
	Cobalt	30.0	+/-30.0	U	30.0	P	11/18/2024	14:35	LB133502
	Copper	20.0	+/-20.0	U	20.0	P	11/18/2024	14:35	LB133502
	Iron	100	+/-100	U	100	P	11/18/2024	14:35	LB133502
	Lead	12.0	+/-12.0	U	12.0	P	11/18/2024	14:35	LB133502
	Magnesium	2000	+/-2000	U	2000	P	11/18/2024	14:35	LB133502
	Manganese	5.73	+/-20.0	J	20.0	P	11/18/2024	14:35	LB133502
	Nickel	40.0	+/-40.0	U	40.0	P	11/18/2024	14:35	LB133502
	Potassium	2000	+/-2000	U	2000	P	11/18/2024	14:35	LB133502
	Selenium	20.0	+/-20.0	U	20.0	P	11/18/2024	14:35	LB133502

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech</u>		SDG No.: <u>P4860</u>							
Contract: <u>CORN02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4860</u>	SAS No.: <u>P4860</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	11/18/2024	14:35	LB133502
	Sodium	2000	+/-2000	U	2000	P	11/18/2024	14:35	LB133502
	Thallium	40.0	+/-40.0	U	40.0	P	11/18/2024	14:35	LB133502
	Vanadium	6.17	+/-40.0	J	40.0	P	11/18/2024	14:35	LB133502
	Zinc	40.0	+/-40.0	U	40.0	P	11/18/2024	14:35	LB133502
CCB03	Aluminum	100	+/-100	U	100	P	11/18/2024	15:39	LB133502
	Antimony	50.0	+/-50.0	U	50.0	P	11/18/2024	15:39	LB133502
	Arsenic	20.0	+/-20.0	U	20.0	P	11/18/2024	15:39	LB133502
	Barium	100	+/-100	U	100	P	11/18/2024	15:39	LB133502
	Beryllium	6.00	+/-6.00	U	6.00	P	11/18/2024	15:39	LB133502
	Cadmium	6.00	+/-6.00	U	6.00	P	11/18/2024	15:39	LB133502
	Calcium	2000	+/-2000	U	2000	P	11/18/2024	15:39	LB133502
	Chromium	10.0	+/-10.0	U	10.0	P	11/18/2024	15:39	LB133502
	Cobalt	30.0	+/-30.0	U	30.0	P	11/18/2024	15:39	LB133502
	Copper	20.0	+/-20.0	U	20.0	P	11/18/2024	15:39	LB133502
	Iron	100	+/-100	U	100	P	11/18/2024	15:39	LB133502
	Lead	12.0	+/-12.0	U	12.0	P	11/18/2024	15:39	LB133502
	Magnesium	2000	+/-2000	U	2000	P	11/18/2024	15:39	LB133502
	Manganese	20.0	+/-20.0	U	20.0	P	11/18/2024	15:39	LB133502
	Nickel	40.0	+/-40.0	U	40.0	P	11/18/2024	15:39	LB133502
	Potassium	2000	+/-2000	U	2000	P	11/18/2024	15:39	LB133502
	Selenium	20.0	+/-20.0	U	20.0	P	11/18/2024	15:39	LB133502
	Silver	10.0	+/-10.0	U	10.0	P	11/18/2024	15:39	LB133502
	Sodium	2000	+/-2000	U	2000	P	11/18/2024	15:39	LB133502
	Thallium	40.0	+/-40.0	U	40.0	P	11/18/2024	15:39	LB133502
	Vanadium	40.0	+/-40.0	U	40.0	P	11/18/2024	15:39	LB133502
	Zinc	40.0	+/-40.0	U	40.0	P	11/18/2024	15:39	LB133502
CCB04	Aluminum	100	+/-100	U	100	P	11/18/2024	16:29	LB133502
	Antimony	50.0	+/-50.0	U	50.0	P	11/18/2024	16:29	LB133502
	Arsenic	20.0	+/-20.0	U	20.0	P	11/18/2024	16:29	LB133502
	Barium	100	+/-100	U	100	P	11/18/2024	16:29	LB133502
	Beryllium	6.00	+/-6.00	U	6.00	P	11/18/2024	16:29	LB133502
	Cadmium	6.00	+/-6.00	U	6.00	P	11/18/2024	16:29	LB133502
	Calcium	2000	+/-2000	U	2000	P	11/18/2024	16:29	LB133502
	Chromium	10.0	+/-10.0	U	10.0	P	11/18/2024	16:29	LB133502
	Cobalt	30.0	+/-30.0	U	30.0	P	11/18/2024	16:29	LB133502
	Copper	20.0	+/-20.0	U	20.0	P	11/18/2024	16:29	LB133502
	Iron	100	+/-100	U	100	P	11/18/2024	16:29	LB133502
	Lead	12.0	+/-12.0	U	12.0	P	11/18/2024	16:29	LB133502

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech</u>		SDG No.: <u>P4860</u>							
Contract: <u>CORN02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4860</u>	SAS No.: <u>P4860</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	11/18/2024	16:29	LB133502
	Manganese	20.0	+/-20.0	U	20.0	P	11/18/2024	16:29	LB133502
	Nickel	40.0	+/-40.0	U	40.0	P	11/18/2024	16:29	LB133502
	Potassium	2000	+/-2000	U	2000	P	11/18/2024	16:29	LB133502
	Selenium	20.0	+/-20.0	U	20.0	P	11/18/2024	16:29	LB133502
	Silver	10.0	+/-10.0	U	10.0	P	11/18/2024	16:29	LB133502
	Sodium	2000	+/-2000	U	2000	P	11/18/2024	16:29	LB133502
	Thallium	40.0	+/-40.0	U	40.0	P	11/18/2024	16:29	LB133502
	Vanadium	40.0	+/-40.0	U	40.0	P	11/18/2024	16:29	LB133502
	Zinc	40.0	+/-40.0	U	40.0	P	11/18/2024	16:29	LB133502
CCB05	Aluminum	100	+/-100	U	100	P	11/18/2024	17:06	LB133502
	Antimony	50.0	+/-50.0	U	50.0	P	11/18/2024	17:06	LB133502
	Arsenic	20.0	+/-20.0	U	20.0	P	11/18/2024	17:06	LB133502
	Barium	100	+/-100	U	100	P	11/18/2024	17:06	LB133502
	Beryllium	6.00	+/-6.00	U	6.00	P	11/18/2024	17:06	LB133502
	Cadmium	6.00	+/-6.00	U	6.00	P	11/18/2024	17:06	LB133502
	Calcium	2000	+/-2000	U	2000	P	11/18/2024	17:06	LB133502
	Chromium	10.0	+/-10.0	U	10.0	P	11/18/2024	17:06	LB133502
	Cobalt	30.0	+/-30.0	U	30.0	P	11/18/2024	17:06	LB133502
	Copper	20.0	+/-20.0	U	20.0	P	11/18/2024	17:06	LB133502
	Iron	100	+/-100	U	100	P	11/18/2024	17:06	LB133502
	Lead	12.0	+/-12.0	U	12.0	P	11/18/2024	17:06	LB133502
	Magnesium	2000	+/-2000	U	2000	P	11/18/2024	17:06	LB133502
	Manganese	20.0	+/-20.0	U	20.0	P	11/18/2024	17:06	LB133502
	Nickel	40.0	+/-40.0	U	40.0	P	11/18/2024	17:06	LB133502
	Potassium	2000	+/-2000	U	2000	P	11/18/2024	17:06	LB133502
	Selenium	20.0	+/-20.0	U	20.0	P	11/18/2024	17:06	LB133502
	Silver	10.0	+/-10.0	U	10.0	P	11/18/2024	17:06	LB133502
	Sodium	2000	+/-2000	U	2000	P	11/18/2024	17:06	LB133502
	Thallium	40.0	+/-40.0	U	40.0	P	11/18/2024	17:06	LB133502
	Vanadium	40.0	+/-40.0	U	40.0	P	11/18/2024	17:06	LB133502
	Zinc	40.0	+/-40.0	U	40.0	P	11/18/2024	17:06	LB133502
CCB06	Aluminum	100	+/-100	U	100	P	11/18/2024	18:08	LB133502
	Antimony	50.0	+/-50.0	U	50.0	P	11/18/2024	18:08	LB133502
	Arsenic	20.0	+/-20.0	U	20.0	P	11/18/2024	18:08	LB133502
	Barium	100	+/-100	U	100	P	11/18/2024	18:08	LB133502
	Beryllium	6.00	+/-6.00	U	6.00	P	11/18/2024	18:08	LB133502
	Cadmium	6.00	+/-6.00	U	6.00	P	11/18/2024	18:08	LB133502
	Calcium	2000	+/-2000	U	2000	P	11/18/2024	18:08	LB133502

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech</u>		SDG No.: <u>P4860</u>							
Contract: <u>CORN02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4860</u>	SAS No.: <u>P4860</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	11/18/2024	18:08	LB133502
	Cobalt	30.0	+/-30.0	U	30.0	P	11/18/2024	18:08	LB133502
	Copper	20.0	+/-20.0	U	20.0	P	11/18/2024	18:08	LB133502
	Iron	100	+/-100	U	100	P	11/18/2024	18:08	LB133502
	Lead	12.0	+/-12.0	U	12.0	P	11/18/2024	18:08	LB133502
	Magnesium	2000	+/-2000	U	2000	P	11/18/2024	18:08	LB133502
	Manganese	20.0	+/-20.0	U	20.0	P	11/18/2024	18:08	LB133502
	Nickel	40.0	+/-40.0	U	40.0	P	11/18/2024	18:08	LB133502
	Potassium	2000	+/-2000	U	2000	P	11/18/2024	18:08	LB133502
	Selenium	20.0	+/-20.0	U	20.0	P	11/18/2024	18:08	LB133502
	Silver	10.0	+/-10.0	U	10.0	P	11/18/2024	18:08	LB133502
	Sodium	2000	+/-2000	U	2000	P	11/18/2024	18:08	LB133502
	Thallium	40.0	+/-40.0	U	40.0	P	11/18/2024	18:08	LB133502
	Vanadium	40.0	+/-40.0	U	40.0	P	11/18/2024	18:08	LB133502
	Zinc	40.0	+/-40.0	U	40.0	P	11/18/2024	18:08	LB133502
CCB07	Aluminum	100	+/-100	U	100	P	11/18/2024	18:58	LB133502
	Antimony	50.0	+/-50.0	U	50.0	P	11/18/2024	18:58	LB133502
	Arsenic	20.0	+/-20.0	U	20.0	P	11/18/2024	18:58	LB133502
	Barium	100	+/-100	U	100	P	11/18/2024	18:58	LB133502
	Beryllium	6.00	+/-6.00	U	6.00	P	11/18/2024	18:58	LB133502
	Cadmium	6.00	+/-6.00	U	6.00	P	11/18/2024	18:58	LB133502
	Calcium	2000	+/-2000	U	2000	P	11/18/2024	18:58	LB133502
	Chromium	10.0	+/-10.0	U	10.0	P	11/18/2024	18:58	LB133502
	Cobalt	30.0	+/-30.0	U	30.0	P	11/18/2024	18:58	LB133502
	Copper	20.0	+/-20.0	U	20.0	P	11/18/2024	18:58	LB133502
	Iron	100	+/-100	U	100	P	11/18/2024	18:58	LB133502
	Lead	12.0	+/-12.0	U	12.0	P	11/18/2024	18:58	LB133502
	Magnesium	2000	+/-2000	U	2000	P	11/18/2024	18:58	LB133502
	Manganese	20.0	+/-20.0	U	20.0	P	11/18/2024	18:58	LB133502
	Nickel	40.0	+/-40.0	U	40.0	P	11/18/2024	18:58	LB133502
	Potassium	2000	+/-2000	U	2000	P	11/18/2024	18:58	LB133502
	Selenium	20.0	+/-20.0	U	20.0	P	11/18/2024	18:58	LB133502
	Silver	10.0	+/-10.0	U	10.0	P	11/18/2024	18:58	LB133502
	Sodium	2000	+/-2000	U	2000	P	11/18/2024	18:58	LB133502
	Thallium	40.0	+/-40.0	U	40.0	P	11/18/2024	18:58	LB133502
	Vanadium	40.0	+/-40.0	U	40.0	P	11/18/2024	18:58	LB133502
	Zinc	40.0	+/-40.0	U	40.0	P	11/18/2024	18:58	LB133502

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech</u>		SDG No.: <u>P4860</u>							
Contract: <u>CORN02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4860</u>	SAS No.: <u>P4860</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	11/19/2024	11:38	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	11:38	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	11:38	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	11:38	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	11:38	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	11:38	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	11:38	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	11:38	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	11:38	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	11:38	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	11:38	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	11:38	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	11:38	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech</u>		SDG No.: <u>P4860</u>							
Contract: <u>CORN02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4860</u>	SAS No.: <u>P4860</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	11/19/2024	12:48	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	12:48	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	12:48	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	12:48	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	12:48	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	12:48	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	12:48	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	12:48	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	12:48	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	12:48	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	12:48	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	12:48	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	12:48	LB133524
CCB02	Aluminum	100	+/-100	U	100	P	11/19/2024	13:38	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	13:38	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	13:38	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	13:38	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	13:38	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	13:38	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	13:38	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	13:38	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	13:38	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	13:38	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech</u>		SDG No.: <u>P4860</u>							
Contract: <u>CORN02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4860</u>	SAS No.: <u>P4860</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	13:38	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	13:38	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	13:38	LB133524
CCB03	Aluminum	100	+/-100	U	100	P	11/19/2024	14:28	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	14:28	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	14:28	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	14:28	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	14:28	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	14:28	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	14:28	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	14:28	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	14:28	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	14:28	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	14:28	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	14:28	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	14:28	LB133524
CCB04	Aluminum	100	+/-100	U	100	P	11/19/2024	15:36	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	15:36	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	15:36	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	15:36	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	15:36	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	15:36	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	15:36	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	15:36	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	15:36	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech</u>		SDG No.: <u>P4860</u>							
Contract: <u>CORN02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4860</u>	SAS No.: <u>P4860</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	15:36	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	15:36	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	15:36	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	15:36	LB133524
CCB05	Aluminum	100	+/-100	U	100	P	11/19/2024	16:22	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	16:22	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Barium	100	+/-100	U	100	P	11/19/2024	16:22	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	16:22	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	16:22	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	16:22	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	16:22	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	16:22	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	16:22	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Manganese	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	16:22	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	16:22	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	16:22	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	16:22	LB133524
CCB06	Aluminum	100	+/-100	U	100	P	11/19/2024	17:24	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	17:24	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	17:24	LB133524
	Barium	22.0	+/-100	J	100	P	11/19/2024	17:24	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	17:24	LB133524
	Cadmium	6.00	+/-6.00	U	6.00	P	11/19/2024	17:24	LB133524
	Calcium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech</u>		SDG No.: <u>P4860</u>							
Contract: <u>CORN02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4860</u>	SAS No.: <u>P4860</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	17:24	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	17:24	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	17:24	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	17:24	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	17:24	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524
	Manganese	6.51	+/-20.0	J	20.0	P	11/19/2024	17:24	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	17:24	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	17:24	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	17:24	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
	Vanadium	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	17:24	LB133524
CCB07	Aluminum	100	+/-100	U	100	P	11/19/2024	18:05	LB133524
	Antimony	50.0	+/-50.0	U	50.0	P	11/19/2024	18:05	LB133524
	Arsenic	20.0	+/-20.0	U	20.0	P	11/19/2024	18:05	LB133524
	Barium	25.8	+/-100	J	100	P	11/19/2024	18:05	LB133524
	Beryllium	6.00	+/-6.00	U	6.00	P	11/19/2024	18:05	LB133524
	Cadmium	0.62	+/-6.00	J	6.00	P	11/19/2024	18:05	LB133524
	Calcium	81.8	+/-2000	J	2000	P	11/19/2024	18:05	LB133524
	Chromium	10.0	+/-10.0	U	10.0	P	11/19/2024	18:05	LB133524
	Cobalt	30.0	+/-30.0	U	30.0	P	11/19/2024	18:05	LB133524
	Copper	20.0	+/-20.0	U	20.0	P	11/19/2024	18:05	LB133524
	Iron	100	+/-100	U	100	P	11/19/2024	18:05	LB133524
	Lead	12.0	+/-12.0	U	12.0	P	11/19/2024	18:05	LB133524
	Magnesium	2000	+/-2000	U	2000	P	11/19/2024	18:05	LB133524
	Manganese	7.68	+/-20.0	J	20.0	P	11/19/2024	18:05	LB133524
	Nickel	40.0	+/-40.0	U	40.0	P	11/19/2024	18:05	LB133524
	Potassium	2000	+/-2000	U	2000	P	11/19/2024	18:05	LB133524
	Selenium	20.0	+/-20.0	U	20.0	P	11/19/2024	18:05	LB133524
	Silver	10.0	+/-10.0	U	10.0	P	11/19/2024	18:05	LB133524
	Sodium	2000	+/-2000	U	2000	P	11/19/2024	18:05	LB133524
	Thallium	40.0	+/-40.0	U	40.0	P	11/19/2024	18:05	LB133524
	Vanadium	6.34	+/-40.0	J	40.0	P	11/19/2024	18:05	LB133524
	Zinc	40.0	+/-40.0	U	40.0	P	11/19/2024	18:05	LB133524