

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech **SDG No.:** P4860
Contract: CORN02 **Lab Code:** CHEM **Case No.:** P4860 **SAS No.:** P4860

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165047BS							
Aluminum	mg/Kg	89.7	95.0		106	80 - 120	P
Antimony	mg/Kg	35.9	36.4		101	80 - 120	P
Arsenic	mg/Kg	35.9	37.2		104	80 - 120	P
Barium	mg/Kg	9.0	9.72		108	80 - 120	P
Beryllium	mg/Kg	9.0	10.1		112	80 - 120	P
Cadmium	mg/Kg	9.0	9.45		105	80 - 120	P
Calcium	mg/Kg	44.8	50.5	J	113	80 - 120	P
Chromium	mg/Kg	17.9	19.3		108	80 - 120	P
Cobalt	mg/Kg	9.0	9.50		106	80 - 120	P
Copper	mg/Kg	13.5	14.8		110	80 - 120	P
Iron	mg/Kg	130	138		106	80 - 120	P
Lead	mg/Kg	44.8	47.2		105	80 - 120	P
Magnesium	mg/Kg	89.7	96.1		107	80 - 120	P
Manganese	mg/Kg	9.0	9.98		111	80 - 120	P
Nickel	mg/Kg	22.4	23.8		106	80 - 120	P
Potassium	mg/Kg	450	440		98	80 - 120	P
Selenium	mg/Kg	89.7	93.6		104	80 - 120	P
Silver	mg/Kg	3.4	3.53		104	80 - 120	P
Sodium	mg/Kg	130	128		98	80 - 120	P
Thallium	mg/Kg	89.7	101		113	80 - 120	P
Vanadium	mg/Kg	13.5	14.6		108	80 - 120	P
Zinc	mg/Kg	9.0	9.92		110	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech **SDG No.:** P4860
Contract: CORN02 **Lab Code:** CHEM **Case No.:** P4860 **SAS No.:** P4860

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165177BS Mercury	mg/Kg	0.26	0.24		94	80 - 120	CV