

Continuing Calibration Report for SequenceID : FC111824AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067806.D

Aliphatic C9-C12	10808106.000	60.000	3.172	6.448	180135.100	165791.359	-8.652
Aliphatic C12-C16	7299351.000	40.000	6.449	9.838	182483.775	170384.422	-7.101
Aliphatic C16-C21	10410399.000	60.000	9.839	13.196	173506.650	169621.553	-2.290
Aliphatic C21-C28	13177298.000	80.000	13.197	16.850	164716.225	157280.060	-4.728
Aliphatic C28-C40	14787957.000	120.000	16.851	21.691	123232.975	121111.336	-1.752
Aliphatic EPH	56483111.000	360.000	3.172	21.691	156897.531	150155.324	-4.490