

Continuing Calibration Report for SequenceID : FC111824AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067818.D

Aliphatic C9-C12	9264897.000	60.000	3.172	6.448	154414.950	165791.359	6.862
Aliphatic C12-C16	6368128.000	40.000	6.449	9.838	159203.200	170384.422	6.562
Aliphatic C16-C21	9355867.000	60.000	9.839	13.196	155931.117	169621.553	8.071
Aliphatic C21-C28	11909221.000	80.000	13.197	16.850	148865.263	157280.060	5.350
Aliphatic C28-C40	13299776.000	120.000	16.851	21.691	110831.467	121111.336	8.488
Aliphatic EPH	50197889.000	360.000	3.172	21.691	139438.581	150155.324	7.137