

Continuing Calibration Report for SequenceID : FD111824AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048777.D

Aromatic EPH	63794665.000	360.000	4.088	18.081	177207.403	173317.266	-2.245
Aromatic C10-C12	8114088.000	40.000	4.088	5.805	202852.200	192017.544	-5.643
Aromatic C12-C16	12005896.000	60.000	5.806	8.411	200098.267	192458.977	-3.969
Aromatic C16-C21	15169934.000	80.000	8.412	12.674	189624.175	185804.816	-2.056
Aromatic C21-C36	28504747.000	180.000	12.675	18.081	158359.706	157231.056	-0.718