

Continuing Calibration Report for SequenceID : FF111924AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015093.D

Aromatic C10-C12	6013023.000	40.000	4.496	6.362	150325.575	136785.903	-9.898
Aromatic C12-C16	8810960.000	60.000	6.363	9.056	146849.333	135022.207	-8.759
Aromatic C16-C21	10961897.000	80.000	9.057	13.369	137023.713	127893.613	-7.139
Aromatic C21-C36	22610114.000	180.000	13.370	18.810	125611.744	121872.475	-3.068
Aromatic EPH	48395994.000	360.000	4.496	18.810	134433.317	127059.175	-5.804