

Continuing Calibration Report for SequenceID : FF111924AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015100.D

Aromatic C10-C12	6030567.000	40.000	4.496	6.362	150764.175	136785.903	-10.219
Aromatic C12-C16	8822475.000	60.000	6.363	9.056	147041.250	135022.207	-8.902
Aromatic C16-C21	10964920.000	80.000	9.057	13.369	137061.500	127893.613	-7.168
Aromatic C21-C36	22282513.000	180.000	13.370	18.810	123791.739	121872.475	-1.575
Aromatic EPH	48100475.000	360.000	4.496	18.810	133612.431	127059.175	-5.158