

Continuing Calibration Report for SequenceID : FG111924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014862.D

Aliphatic C9-C12	7716226.000	60.000	3.218	6.839	128603.767	138200.947	6.944
Aliphatic C12-C16	5227929.000	40.000	6.840	10.283	130698.225	140164.944	6.754
Aliphatic C16-C21	7732718.000	60.000	10.284	13.654	128878.633	134419.042	4.122
Aliphatic C21-C28	9435176.000	80.000	13.655	17.319	117939.700	123528.952	4.525
Aliphatic C28-C40	10421500.000	120.000	17.320	22.248	86845.833	103690.651	16.245
Aliphatic EPH	40533549.000	360.000	3.218	22.248	112593.192	123024.976	8.479