

Continuing Calibration Report for SequenceID : FG111924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014871.D

Aliphatic C9-C12	7754881.000	60.000	3.218	6.839	129248.017	138200.947	6.478
Aliphatic C12-C16	5244895.000	40.000	6.840	10.283	131122.375	140164.944	6.451
Aliphatic C16-C21	7664939.000	60.000	10.284	13.654	127748.983	134419.042	4.962
Aliphatic C21-C28	9297761.000	80.000	13.655	17.319	116222.013	123528.952	5.915
Aliphatic C28-C40	10532125.000	120.000	17.320	22.248	87767.708	103690.651	15.356
Aliphatic EPH	40494601.000	360.000	3.218	22.248	112485.003	123024.976	8.567