

SOIL EPH SURROGATE RECOVERY

Lab Name: CHEMTECH Contract: CORN02
Lab Code: CHEM CASE No.: P4860 SAS No.: P4860 SDG No.: P4860
Run Number: FC111824AL

Client SAMPLE NO.	1-chlorooctadecane (SURR)			TOT OUT
EX-4-TPH-6MS	78			0
EX-4-TPH-6MSD	77			0
PH2-BOT-002	76			0
PH2-BOT-003	58			0
PH2-BOT-004	52			0
PH2-BOT-009	55			0
PB165032BL	75			0
PB165032BS	65			0
PB165032BSD	64			0

QC LIMITS

1-chlorooctadecane (SURR)

(40-140)

Column to be used to flag recovery values
* Values outside of contract required QC Limits
D Surrogate diluted out

SOIL EPH SURROGATE RECOVERY

Lab Name: CHEMTECH Contract: CORN02
Lab Code: CHEM CASE No.: P4860 SAS No.: P4860 SDG No.: P4860
Run Number: FD111824AR

Client SAMPLE NO.	2-Bromonaphthalene (SURR)	2-Fluorobiphenyl (SURR)	ortho-Terphenyl (SURR)	TOT OUT
EX-4-TPH-6MS	102	101	78	0
EX-4-TPH-6MSD	102	101	77	0
DUP-01	105	108	57	0
PH2-BOT-001	115	118	67	0
PH2-BOT-002	94	81	77	0
PH2-BOT-003	98	99	44	0
PH2-BOT-004	116	112	64	0
PH2-BOT-009	90	87	55	0
PB165032BL	104	102	96	0
PB165032BS	92	84	79	0
PB165032BSD	94	85	80	0

QC LIMITS

2-Bromonaphthalene (SURR) (40-140)
2-Fluorobiphenyl (SURR) (40-140)
ortho-Terphenyl (SURR) (40-140)

Column to be used to flag recovery values
* Values outside of contract required QC Limits
D Surrogate diluted out

SOIL EPH SURROGATE RECOVERY

Lab Name: CHEMTECH Contract: CORN02
Lab Code: CHEM CASE No.: P4860 SAS No.: P4860 SDG No.: P4860
Run Number: FF111924AR

Client SAMPLE NO.	2-Bromonaphthalene (SURR)	2-Flurobiphenyl (SURR)	ortho-Terphenyl (SURR)	TOT OUT
PH2-BOT-008	102	91	80	0
PH2-BOT-007	104	97	49	0
PH2-BOT-006	104	100	57	0
PH2-BOT-005	105	96	57	0

QC LIMITS

2-Bromonaphthalene (SURR) (40-140)
2-Flurobiphenyl (SURR) (40-140)
ortho-Terphenyl (SURR) (40-140)

Column to be used to flag recovery values
* Values outside of contract required QC Limits
D Surrogate diluted out

SOIL EPH SURROGATE RECOVERY

Lab Name: CHEMTECH Contract: CORN02
Lab Code: CHEM CASE No.: P4860 SAS No.: P4860 SDG No.: P4860
Run Number: FG111924AL

Client SAMPLE NO.	1-chlorooctadecane (SURR)			TOT OUT
DUP-01	68			0
PH2-BOT-001	85			0
PH2-BOT-008	81			0
PH2-BOT-007	63			0
PH2-BOT-006	69			0
PH2-BOT-005	68			0

QC LIMITS

1-chlorooctadecane (SURR)

(40-140)

Column to be used to flag recovery values
* Values outside of contract required QC Limits
D Surrogate diluted out

SOIL EPH SURROGATE RECOVERY

QC LIMITS

1-chlorooctadecane (SURR)

(40-140)

Column to be used to flag recovery values
* Values outside of contract required QC Limits
D Surrogate diluted out