

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4860

Client: Tetra Tech

Analytical Method: 8081B

DataFile : PL093148.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: P4860-08MSD	PH2-BOT-007MSD											
	alpha-BHC	16.98	0	15.9	ug/kg	94		1		30 (60)	150 (144)	30 (20)
	beta-BHC	16.98	0	16.6	ug/kg	98		1		30 (54)	150 (143)	30 (20)
	delta-BHC	16.98	0	13.3	ug/kg	78		1		30 (47)	150 (144)	30 (20)
	gamma-BHC (Lindane)	16.98	0	15.6	ug/kg	92		2		30 (61)	150 (140)	30 (20)
	Heptachlor	16.98	0	16.7	ug/kg	98		2		30 (63)	150 (135)	30 (20)
	Aldrin	16.98	0	16.2	ug/kg	95		2		30 (49)	150 (139)	30 (20)
	Heptachlor epoxide	16.98	0	16.3	ug/kg	96		1		30 (32)	150 (180)	30 (20)
	Endosulfan I	16.98	0	15.1	ug/kg	89		5		30 (56)	150 (142)	30 (20)
	Dieldrin	16.98	0	17.0	ug/kg	100		1		30 (47)	150 (161)	30 (20)
	4,4'-DDE	16.98	0.476	17.8	ug/kg	102		0		30 (55)	150 (136)	30 (20)
	Endrin	16.98	0	18.3	ug/kg	108		3		30 (57)	150 (139)	30 (20)
	Endosulfan II	16.98	0	16.8	ug/kg	99		0		30 (40)	150 (163)	30 (20)
	4,4'-DDD	16.98	1	18.0	ug/kg	100		2		30 (37)	150 (192)	30 (20)
	Endosulfan sulfate	16.98	0	16.0	ug/kg	94		0		30 (62)	150 (139)	30 (20)
	4,4'-DDT	16.98	1.7	17.6	ug/kg	94		2		30 (51)	150 (146)	30 (20)
	Methoxychlor	16.98	0	16.6	ug/kg	98		1		30 (54)	150 (136)	30 (20)
	Endrin ketone	16.98	0	16.0	ug/kg	94		1		30 (60)	150 (129)	30 (20)
	Endrin aldehyde	16.98	0	15.5	ug/kg	91		1		30 (59)	150 (132)	30 (20)
	alpha-Chlordane	16.98	0	17.0	ug/kg	100		2		30 (30)	150 (192)	30 (20)
	gamma-Chlordane	16.98	0	16.9	ug/kg	100		1		30 (44)	150 (175)	30 (20)