

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111324

Review By	yogesh	Review On	11/14/2024 3:08:53 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:17 AM
SubDirectory	BF111324	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6573 SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6624 S12327,10ul/1000ul sample SP6559

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140331.D	13 Nov 2024 08:35		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF140332.D	13 Nov 2024 09:01		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF140333.D	13 Nov 2024 09:27	Compound#41,54,77 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF140334.D	13 Nov 2024 09:53		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF140335.D	13 Nov 2024 10:29		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF140336.D	13 Nov 2024 10:55	Injection Error	RC/JU	Not Ok
7	SSTDICC050	SSTDICC050	BF140337.D	13 Nov 2024 11:21		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF140338.D	13 Nov 2024 11:47		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF140339.D	13 Nov 2024 12:13	Compound#09 removed from 80 ppm	RC/JU	Ok,M
10	SSTDICCC040	SSTDICCC040	BF140340.D	13 Nov 2024 12:48		RC/JU	Ok
11	SSTDICV040	ICVBF111324	BF140341.D	13 Nov 2024 13:19		RC/JU	Ok
12	PB164401BL	PB164401BL	BF140342.D	13 Nov 2024 13:45		RC/JU	Ok
13	P4495-11	PT-BNA-SOIL	BF140343.D	13 Nov 2024 15:25	PT Sample, Need 5X Dilution	RC/JU	Dilution
14	P4495-11DL	PT-BNA-SOILDL	BF140344.D	13 Nov 2024 16:02		RC/JU	Ok
15	P3845-03	PT-BN-WP	BF140345.D	13 Nov 2024 16:40	PT Sample Analyzed for confirmation - Need 5X Dilution.	RC/JU	Dilution
16	P3845-03DL	PT-BN-WPDL	BF140346.D	13 Nov 2024 17:06		RC/JU	Ok
17	P3845-06	PT-ACIDS-WP	BF140347.D	13 Nov 2024 17:32	PT Sample Analyzed for confirmation - Need 5X Dilution.	RC/JU	Dilution

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111324

Review By	yogesh	Review On	11/14/2024 3:08:53 AM		
Supervise By	mohammad	Supervise On	11/15/2024 5:49:17 AM		
SubDirectory	BF111324	HP Acquire Method	BNA_F	HP Processing Method	Bf111324
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12327,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P3845-06DL	PT-ACIDS-WPDL	BF140348.D	13 Nov 2024 17:58	Dilution not matching (Low results)	RC/JU	Not Ok
19	SP6681	SP6681	BF140349.D	13 Nov 2024 18:39	8270-625.1 SPIKE SOLUTION	RC/JU	Ok,NS

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111824

Review By	yogesh	Review On	11/19/2024 3:20:55 AM
Supervise By	mohammad	Supervise On	11/19/2024 3:30:54 AM
SubDirectory	BF111824	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140442.D	18 Nov 2024 09:18		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140443.D	18 Nov 2024 10:13		RC/JU	Ok,M
3	PB165029BL	PB165029BL	BF140444.D	18 Nov 2024 10:39		RC/JU	Ok
4	PB165029BS	PB165029BS	BF140445.D	18 Nov 2024 11:05		RC/JU	Ok,M
5	PB164846BS	PB164846BS	BF140446.D	18 Nov 2024 11:32		RC/JU	Ok,M
6	PB164846BSD	PB164846BSD	BF140447.D	18 Nov 2024 11:58		RC/JU	Ok,M
7	PB164940BL	PB164940BL	BF140448.D	18 Nov 2024 12:24		RC/JU	Ok
8	PB164940BS	PB164940BS	BF140449.D	18 Nov 2024 12:50		RC/JU	Ok,M
9	PB164965BL	PB164965BL	BF140450.D	18 Nov 2024 13:16		RC/JU	Ok
10	PB164965BS	PB164965BS	BF140451.D	18 Nov 2024 13:43		RC/JU	Ok,M
11	PB164969BS	PB164969BS	BF140452.D	18 Nov 2024 14:19		RC/JU	Ok,M
12	PB164969BL	PB164969BL	BF140453.D	18 Nov 2024 14:45		RC/JU	Ok
13	PB164880TB	PB164880TB	BF140454.D	18 Nov 2024 15:11		RC/JU	Ok
14	DFTPP	DFTPP	BF140455.D	18 Nov 2024 15:48		RC/JU	Ok
15	SSTDCCC040	SSTDCCC040	BF140456.D	18 Nov 2024 16:14		RC/JU	Ok,M
16	PB164944TB	PB164944TB	BF140457.D	18 Nov 2024 16:40		RC/JU	Ok
17	P4821-04RE	BP-F-24RE	BF140458.D	18 Nov 2024 17:13	Surrogate Fail	RC/JU	Confirms
18	P4861-01	WC-11-A-202411	BF140459.D	18 Nov 2024 17:40		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111824

Review By	yogesh	Review On	11/19/2024 3:20:55 AM
Supervise By	mohammad	Supervise On	11/19/2024 3:30:54 AM
SubDirectory	BF111824	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P4871-01	WC-10-A-202411	BF140460.D	18 Nov 2024 18:06		RC/JU	Ok
20	P4848-02	TP-1	BF140461.D	18 Nov 2024 18:32		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112124

Review By	yogesh	Review On	11/22/2024 3:53:36 AM
Supervise By	mohammad	Supervise On	11/27/2024 6:04:00 AM
SubDirectory	BF112124	HP Acquire Method	BNA_F
		HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12329,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140526.D	21 Nov 2024 10:17		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140527.D	21 Nov 2024 10:44	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF140528.D	21 Nov 2024 11:13		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BF140529.D	21 Nov 2024 11:39	Compound#9,32,41,54,56,65,70 removed from 5 ppm	RC/JU	Ok
5	SSTDICC010	SSTDICC010	BF140530.D	21 Nov 2024 12:05		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BF140531.D	21 Nov 2024 12:32	Compound#32,41,54 Kept on LR	RC/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BF140532.D	21 Nov 2024 12:58	Calibration failed for Benzidine	RC/JU	Ok,M
8	SSTDICC050	SSTDICC050	BF140533.D	21 Nov 2024 13:25	The Calibration is Good For 8270 DOD Except com#77 and good for 625.1 Method Except com#77	RC/JU	Ok
9	SSTDICC060	SSTDICC060	BF140534.D	21 Nov 2024 13:51		RC/JU	Ok
10	SSTDICC080	SSTDICC080	BF140535.D	21 Nov 2024 14:18	Compound#9 removed from 80 ppm	RC/JU	Ok
11	SSTDICV040	ICVBF112124	BF140536.D	21 Nov 2024 15:07		RC/JU	Ok,M
12	PB165144BL	PB165144BL	BF140537.D	21 Nov 2024 15:34		RC/JU	Ok
13	DFTPP	DFTPP	BF140538.D	21 Nov 2024 16:27		RC/JU	Ok
14	SSTDCCC040	SSTDCCC040	BF140539.D	21 Nov 2024 16:54		RC/JU	Ok,M
15	PB165060TB	PB165060TB	BF140540.D	21 Nov 2024 17:20		RC/JU	Ok
16	P4887-06	MH-760	BF140541.D	21 Nov 2024 17:55		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112124

Review By	yogesh	Review On	11/22/2024 3:53:36 AM		
Supervise By	mohammad	Supervise On	11/27/2024 6:04:00 AM		
SubDirectory	BF112124	HP Acquire Method	BNA_F	HP Processing Method	bf112124
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12329,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

17	P4887-02	MH-739	BF140542.D	21 Nov 2024 18:21		RC/JU	Ok
18	P4870-16	TP-15	BF140543.D	21 Nov 2024 18:48		RC/JU	Ok
19	P4870-15	MH-736	BF140544.D	21 Nov 2024 19:14		RC/JU	Ok
20	P4870-14	MH-735	BF140545.D	21 Nov 2024 19:40		RC/JU	Ok
21	P4870-13	TP-1	BF140546.D	21 Nov 2024 20:07		RC/JU	Ok
22	P4929-02	ARS520	BF140547.D	21 Nov 2024 20:33		RC/JU	Ok
23	P4924-04	MH-4	BF140548.D	21 Nov 2024 20:59		RC/JU	Ok
24	P4921-01	WC-11-A-202411	BF140549.D	21 Nov 2024 21:25		RC/JU	Ok
25	P4916-12	TP-3-WC	BF140550.D	21 Nov 2024 21:51		RC/JU	Ok
26	P4916-08	TP-2-WC	BF140551.D	21 Nov 2024 22:18		RC/JU	Ok
27	P4916-04	TP-1-WC	BF140552.D	21 Nov 2024 22:44		RC/JU	Ok
28	P4887-01	MH-739	BF140553.D	21 Nov 2024 23:10		RC/JU	Ok
29	P4916-09	TP-3-WC	BF140554.D	21 Nov 2024 23:36		RC/JU	Ok
30	P4916-09MS	TP-3-WCMS	BF140555.D	22 Nov 2024 00:02		RC/JU	Ok,M
31	P4916-09MSD	TP-3-WCMSD	BF140556.D	22 Nov 2024 00:28		RC/JU	Ok,M
32	P4916-01	TP-1-WC	BF140557.D	22 Nov 2024 00:54		RC/JU	Ok
33	P4916-05RE	TP-2-WCRE	BF140558.D	22 Nov 2024 01:21	Internal Standard Fail	RC/JU	Confirms
34	P4887-05RE	MH-760RE	BF140559.D	22 Nov 2024 01:47	Internal Standard Fail	RC/JU	Confirms
35	P4924-01	MH-4	BF140560.D	22 Nov 2024 02:13	Internal Standard Fail	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112124

Review By	yogesh	Review On	11/22/2024 3:53:36 AM		
Supervise By	mohammad	Supervise On	11/27/2024 6:04:00 AM		
SubDirectory	BF112124	HP Acquire Method	BNA_F	HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12329,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

36	P4936-01	PL-01-11202024	BF140561.D	22 Nov 2024 02:39	Internal Standard Fail	RC/JU	Ok,M
37	P4929-01	ARS520	BF140562.D	22 Nov 2024 03:06	Internal Standard Fail	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112224

Review By	Jagrut	Review On	11/25/2024 11:47:00 AM
Supervise By	mohammad	Supervise On	11/27/2024 6:04:05 AM
SubDirectory	BF112224	HP Acquire Method	BNA_F
		HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12329,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140563.D	22 Nov 2024 08:54		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140564.D	22 Nov 2024 09:20		RC/JU	Ok
3	PB165129BL	PB165129BL	BF140565.D	22 Nov 2024 09:46		RC/JU	Ok
4	PB165129BS	PB165129BS	BF140566.D	22 Nov 2024 10:12		RC/JU	Ok,M
5	P4910-08	MH-759	BF140567.D	22 Nov 2024 10:44		RC/JU	Ok
6	P4910-04	MH-COTTAGE	BF140568.D	22 Nov 2024 11:10		RC/JU	Ok
7	P4893-08	MH-762	BF140569.D	22 Nov 2024 11:36		RC/JU	Ok
8	P4893-04	MH-763	BF140570.D	22 Nov 2024 12:02		RC/JU	Ok
9	P4925-08	MH-741	BF140571.D	22 Nov 2024 12:28		RC/JU	Ok
10	P4925-04	MH-741	BF140572.D	22 Nov 2024 12:55		RC/JU	Ok
11	P4923-06	72-11991	BF140573.D	22 Nov 2024 13:21		RC/JU	Ok
12	P4925-01	MH-741	BF140574.D	22 Nov 2024 13:47		RC/JU	Ok
13	P4938-01	MH-732	BF140575.D	22 Nov 2024 14:13		RC/JU	Ok
14	P4938-01MS	MH-732MS	BF140576.D	22 Nov 2024 14:40		RC/JU	Ok,M
15	P4938-01MSD	MH-732MSD	BF140577.D	22 Nov 2024 15:06		RC/JU	Ok,M
16	P4860-08	PH2-BOT-007	BF140578.D	22 Nov 2024 15:32		RC/JU	Ok
17	P4860-05	PH2-BOT-004	BF140579.D	22 Nov 2024 15:58		RC/JU	Ok
18	P4924-01	MH-4	BF140580.D	22 Nov 2024 16:24		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112224

Review By	Jagrut	Review On	11/25/2024 11:47:00 AM		
Supervise By	mohammad	Supervise On	11/27/2024 6:04:05 AM		
SubDirectory	BF112224	HP Acquire Method	BNA_F	HP Processing Method	bf112124
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12329,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4948-01	337	BF140581.D	22 Nov 2024 16:50		RC/JU	Ok,M
20	P4949-03	CONTAINMENT-STON	BF140582.D	22 Nov 2024 17:16	Internal Standard Fail	RC/JU	Ok
21	P4948-03	72-11944	BF140583.D	22 Nov 2024 17:43	Internal Standard Fail	RC/JU	Ok,M
22	P4822-04	SOIL-2	BF140584.D	22 Nov 2024 18:09		RC/JU	Ok
23	P4822-02	SOIL-1	BF140585.D	22 Nov 2024 18:35		RC/JU	Ok,M
24	P4925-05	MH-758	BF140586.D	22 Nov 2024 19:01		RC/JU	Ok
25	P4938-05	MH-734	BF140587.D	22 Nov 2024 19:27		RC/JU	Ok
26	P4860-02	PH2-BOT-001	BF140588.D	22 Nov 2024 19:53		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112524

Review By	yogesh	Review On	11/27/2024 5:27:46 AM
Supervise By	mohammad	Supervise On	11/27/2024 6:04:35 AM
SubDirectory	BF112524	HP Acquire Method	BNA_F
		HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12330,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140589.D	25 Nov 2024 09:07		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140590.D	25 Nov 2024 09:33		RC/JU	Ok,M
3	PB165123TB	PB165123TB	BF140591.D	25 Nov 2024 09:59		RC/JU	Ok
4	PB165155BS	PB165155BS	BF140592.D	25 Nov 2024 10:25		RC/JU	Ok,M
5	PB165155BL	PB165155BL	BF140593.D	25 Nov 2024 10:51		RC/JU	Ok
6	PB165086BS	PB165086BS	BF140594.D	25 Nov 2024 11:17		RC/JU	Ok,M
7	PB165111TB	PB165111TB	BF140595.D	25 Nov 2024 11:43		RC/JU	Ok
8	PB165052BS	PB165052BS	BF140596.D	25 Nov 2024 12:09		RC/JU	Ok,M
9	PB164986TB	PB164986TB	BF140597.D	25 Nov 2024 12:36		RC/JU	Ok
10	PB165152BS	PB165152BS	BF140598.D	25 Nov 2024 13:02		RC/JU	Ok,M
11	PB164886TB	PB164886TB	BF140599.D	25 Nov 2024 13:27		RC/JU	Ok
12	PB165152BSD	PB165152BSD	BF140600.D	25 Nov 2024 13:54		RC/JU	Ok,M
13	PB165152BL	PB165152BL	BF140601.D	25 Nov 2024 14:28		RC/JU	Ok
14	PB165185BS	PB165185BS	BF140602.D	25 Nov 2024 14:54		RC/JU	Ok,M
15	DFTPP	DFTPP	BF140603.D	25 Nov 2024 15:23		RC/JU	Ok
16	SSTDCCC040	SSTDCCC040	BF140604.D	25 Nov 2024 15:49		RC/JU	Ok,M
17	PB165052BL	PB165052BL	BF140605.D	25 Nov 2024 16:17		RC/JU	Ok
18	P4947-01	A3988	BF140606.D	25 Nov 2024 16:49		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112524

Review By	yogesh	Review On	11/27/2024 5:27:46 AM
Supervise By	mohammad	Supervise On	11/27/2024 6:04:35 AM
SubDirectory	BF112524	HP Acquire Method	BNA_F
		HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12330,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P4892-04	WB-310-SW	BF140607.D	25 Nov 2024 17:15		RC/JU	Ok
20	P4860-09	PH2-BOT-006	BF140608.D	25 Nov 2024 17:41		RC/JU	Ok
21	P4860-01	DUP-01	BF140609.D	25 Nov 2024 18:07		RC/JU	Ok
22	P4960-01	B1	BF140610.D	25 Nov 2024 18:34		RC/JU	Ok
23	P4960-05	SW3	BF140611.D	25 Nov 2024 19:00		RC/JU	Ok
24	P4892-03	WB-310-BOT	BF140612.D	25 Nov 2024 19:26		RC/JU	Ok
25	P4892-03MS	WB-310-BOTMS	BF140613.D	25 Nov 2024 19:52		RC/JU	Ok,M
26	P4892-03MSD	WB-310-BOTMSD	BF140614.D	25 Nov 2024 20:18		RC/JU	Ok,M
27	P4860-06	PH2-BOT-009	BF140615.D	25 Nov 2024 20:45		RC/JU	Ok
28	P4860-09MS	PH2-BOT-006MS	BF140616.D	25 Nov 2024 21:11	MSD Not Ok	RC/JU	Not Ok
29	P4860-09MSD	PH2-BOT-006MSD	BF140617.D	25 Nov 2024 21:37	Internal Standard Fail	RC/JU	Not Ok
30	P4960-03	SW1	BF140618.D	25 Nov 2024 22:03	Internal Standard Fail	RC/JU	ReRun
31	P4860-10	PH2-BOT-005	BF140619.D	25 Nov 2024 22:29	Internal Standard Fail	RC/JU	ReRun
32	P4860-04	PH2-BOT-003	BF140620.D	25 Nov 2024 22:55		RC/JU	Ok
33	P4860-07	PH2-BOT-008	BF140621.D	25 Nov 2024 23:22		RC/JU	Ok
34	P4860-03	PH2-BOT-002	BF140622.D	25 Nov 2024 23:48		RC/JU	Ok
35	P4951-01	AU-05-112124	BF140623.D	26 Nov 2024 00:14	Internal Standard Fail	RC/JU	ReRun
36	P4954-01	TR-05-112124	BF140624.D	26 Nov 2024 00:40	Internal Standard Fail	RC/JU	Ok,M
37	P4954-01MS	TR-05-112124MS	BF140625.D	26 Nov 2024 01:06	Internal Standard Fail	RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112524

Review By	yogesh	Review On	11/27/2024 5:27:46 AM		
Supervise By	mohammad	Supervise On	11/27/2024 6:04:35 AM		
SubDirectory	BF112524	HP Acquire Method	BNA_F	HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12330,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

38	P4954-01MSD	TR-05-112124MSD	BF140626.D	26 Nov 2024 01:32	Internal Standard Fail	RC/JU	Ok,M
39	P4954-03	TR-06-112124	BF140627.D	26 Nov 2024 01:58	Internal Standard Fail	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112624

Review By	yogesh	Review On	11/27/2024 5:33:18 AM
Supervise By	mohammad	Supervise On	11/27/2024 6:05:09 AM
SubDirectory	BF112624	HP Acquire Method	BNA_F
		HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12330,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140628.D	26 Nov 2024 09:09		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140629.D	26 Nov 2024 09:35		RC/JU	Ok,M
3	PB165203BL	PB165203BL	BF140630.D	26 Nov 2024 10:01		RC/JU	Ok
4	PB165203BS	PB165203BS	BF140631.D	26 Nov 2024 10:27		RC/JU	Ok,M
5	PB165185BL	PB165185BL	BF140632.D	26 Nov 2024 10:53		RC/JU	Ok
6	PB165203BSD	PB165203BSD	BF140633.D	26 Nov 2024 11:20		RC/JU	Ok,M
7	P4960-04	SW2	BF140634.D	26 Nov 2024 11:52		RC/JU	Ok
8	P4960-06	SW4	BF140635.D	26 Nov 2024 12:18		RC/JU	Ok
9	P4960-02	B2	BF140636.D	26 Nov 2024 12:44		RC/JU	Ok
10	P5000-01	MH-745	BF140637.D	26 Nov 2024 13:10		RC/JU	Ok
11	P4960-03	SW1	BF140638.D	26 Nov 2024 13:36		RC/JU	Ok
12	P4860-09MS	PH2-BOT-006MS	BF140639.D	26 Nov 2024 14:03		RC/JU	Ok,M
13	P4860-09MSD	PH2-BOT-006MSD	BF140640.D	26 Nov 2024 14:29		RC/JU	Ok,M
14	P4860-10	PH2-BOT-005	BF140641.D	26 Nov 2024 14:55		RC/JU	Ok
15	P4951-01RE	AU-05-112124RE	BF140642.D	26 Nov 2024 15:22	Internal Standard Fail	RC/JU	Confirms
16	P4985-01	MH-756-WC	BF140643.D	26 Nov 2024 15:48	Internal Standard Fail	RC/JU	Ok
17	P4985-01MS	MH-756-WCMS	BF140644.D	26 Nov 2024 16:14	Internal Standard Fail	RC/JU	Ok,M
18	P4985-01MSD	MH-756-WCMSD	BF140645.D	26 Nov 2024 16:41	Internal standard fail	RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112624

Review By	yogesh	Review On	11/27/2024 5:33:18 AM
Supervise By	mohammad	Supervise On	11/27/2024 6:05:09 AM
SubDirectory	BF112624	HP Acquire Method	BNA_F
		HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12330,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P4938-04	MH-732	BF140646.D	26 Nov 2024 17:07	Internal Standard Fail	RC/JU	ReRun
20	P4938-04MS	MH-732MS	BF140647.D	26 Nov 2024 17:33	Internal Standard Fail	RC/JU	Not Ok
21	P4938-04MSD	MH-732MSD	BF140648.D	26 Nov 2024 18:00	MS is Not Ok	RC/JU	Not Ok
22	P4985-04	MH-756-WC	BF140649.D	26 Nov 2024 18:26	Internal Standard Fail	RC/JU	ReRun
23	P4938-08	MH-734	BF140650.D	26 Nov 2024 18:52	Internal Standard Fail	RC/JU	ReRun
24	P4985-05	MH-740-WC	BF140651.D	26 Nov 2024 19:19	Internal Standard Fail	RC/JU	ReRun
25	P5000-05	MH-733	BF140652.D	26 Nov 2024 19:45	Internal Standard Fail	RC/JU	ReRun
26	P4971-01	CL-02-112224	BF140653.D	26 Nov 2024 20:11		RC/JU	Ok,M

M : Manual Integration