

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4860 SAS No.: P4860 SDG NO.: P4860

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/22/2024

Lab File ID: BF140564.D Time Analyzed: 09:20

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	98979	6.869	364482	8.15	206333	9.91
UPPER LIMIT	197958	7.369	728964	8.651	412666	10.41
LOWER LIMIT	49489.5	6.369	182241	7.651	103167	9.41
EPA SAMPLE NO.						
01 PH2-BOT-004	90927	6.87	332795	8.15	173346	9.90
02 PH2-BOT-007	93283	6.87	346410	8.15	184115	9.90
03 PH2-BOT-001	91972	6.87	309675	8.15	141425	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.