

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CORN02
 Lab Code: CHEM Case No.: P4860 SAS No.: P4860 SDG No.: P4860
 Instrument ID: MSVOA_D Calibration Date(s): 11/15/2024 11/15/2024
 Heated Purge: (Y/N) Y Calibration Time(s): 10:04 12:18
 GC Column: RTX-VMS ID: 0.18 (mm)

LAB FILE ID:		RRF005 = VD080020.D		RRF010 = VD080021.D		RRF020 = VD080022.D			
		RRF050 = VD080023.D		RRF100 = VD080024.D		RRF150 = VD080025.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Dichlorodifluoromethane	0.465	0.388	0.364	0.396	0.396	0.399	0.401	8.4	
Chloromethane	0.330	0.348	0.279	0.276	0.281	0.257	0.295	12.1	
Vinyl Chloride	0.377	0.390	0.312	0.322	0.330	0.290	0.337	11.5	
Bromomethane	0.451	0.498	0.345	0.305	0.347	0.296	0.374	22	
Chloroethane	0.254	0.309	0.228	0.208	0.251	0.195	0.241	16.9	
Trichlorofluoromethane	0.935	0.919	0.763	0.766	0.824	0.758	0.827	9.8	
1,1,2-Trichlorotrifluoroethane	0.571	0.563	0.486	0.481	0.498	0.503	0.517	7.6	
1,1-Dichloroethene	0.535	0.540	0.494	0.491	0.519	0.529	0.518	4.1	
Acetone	0.101	0.116	0.091	0.106	0.132	0.124	0.112	13.6	
Carbon Disulfide	1.724	1.793	1.552	1.554	1.697	1.690	1.668	5.8	
Methyl tert-butyl Ether	0.877	1.057	1.011	1.088	1.156	1.153	1.057	9.9	
Methyl Acetate	0.270	0.251	0.248	0.214	0.255	0.249	0.248	7.5	
Methylene Chloride	0.643	0.628	0.612	0.556	0.581	0.575	0.599	5.6	
trans-1,2-Dichloroethene	0.589	0.579	0.536	0.578	0.579	0.576	0.573	3.3	
1,1-Dichloroethane	1.015	1.040	0.961	0.973	0.974	0.972	0.989	3.2	
Cyclohexane	0.945	0.869	0.780	0.797	0.811	0.828	0.838	7.2	
2-Butanone	0.135	0.141	0.139	0.154	0.155	0.152	0.146	5.8	
Carbon Tetrachloride	0.521	0.515	0.482	0.491	0.483	0.490	0.497	3.4	
cis-1,2-Dichloroethene	0.665	0.649	0.620	0.656	0.675	0.679	0.657	3.2	
Bromochloromethane	0.506	0.495	0.416	0.406	0.402	0.415	0.440	10.7	
Chloroform	1.057	1.143	1.012	1.019	1.006	1.015	1.042	5	
1,1,1-Trichloroethane	0.970	0.930	0.851	0.869	0.864	0.867	0.892	5.3	
Methylcyclohexane	0.517	0.515	0.476	0.556	0.583	0.604	0.542	8.9	
Benzene	1.391	1.430	1.351	1.436	1.439	1.463	1.418	2.8	
1,2-Dichloroethane	0.376	0.382	0.350	0.368	0.367	0.359	0.367	3.1	
Trichloroethene	0.378	0.389	0.341	0.361	0.361	0.372	0.367	4.5	
1,2-Dichloropropane	0.333	0.345	0.329	0.352	0.342	0.344	0.341	2.5	
Bromodichloromethane	0.513	0.504	0.472	0.499	0.492	0.497	0.496	2.8	
4-Methyl-2-Pentanone	0.160	0.164	0.172	0.193	0.197	0.193	0.180	9.1	
Toluene	0.814	0.864	0.823	0.917	0.927	0.945	0.881	6.4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CORN02
 Lab Code: CHEM Case No.: P4860 SAS No.: P4860 SDG No.: P4860
 Instrument ID: MSVOA_D Calibration Date(s): 11/15/2024 11/15/2024
 Heated Purge: (Y/N) Y Calibration Time(s): 10:04 12:18
 GC Column: RTX-VMS ID: 0.18 (mm)

LAB FILE ID:		RRF005 = VD080020.D		RRF010 = VD080021.D		RRF020 = VD080022.D			
		RRF050 = VD080023.D		RRF100 = VD080024.D		RRF150 = VD080025.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.401	0.440	0.420	0.459	0.465	0.471	0.442	6.2	
cis-1,3-Dichloropropene	0.490	0.501	0.506	0.542	0.549	0.555	0.524	5.3	
1,1,2-Trichloroethane	0.256	0.276	0.250	0.267	0.262	0.263	0.262	3.4	
2-Hexanone	0.108	0.124	0.122	0.147	0.148	0.145	0.132	12.5	
Dibromochloromethane	0.317	0.348	0.320	0.349	0.343	0.349	0.338	4.5	
1,2-Dibromoethane	0.234	0.251	0.237	0.252	0.256	0.255	0.247	3.9	
Tetrachloroethene	0.386	0.365	0.337	0.342	0.352	0.355	0.356	5	
Chlorobenzene	1.138	1.102	1.024	1.070	1.095	1.116	1.091	3.6	
Ethyl Benzene	1.740	1.714	1.693	1.866	1.965	2.002	1.830	7.3	
m/p-Xylenes	0.647	0.688	0.683	0.727	0.766	0.780	0.715	7.2	
o-Xylene	0.542	0.595	0.604	0.670	0.720	0.726	0.643	11.5	
Styrene	0.969	1.068	1.105	1.212	1.259	1.271	1.148	10.5	
Bromoform	0.240	0.222	0.215	0.223	0.222	0.221	0.224	3.8	
Isopropylbenzene	3.045	3.093	2.974	3.271	3.463	3.501	3.224	6.9	
1,1,2,2-Tetrachloroethane	0.661	0.656	0.585	0.608	0.612	0.597	0.620	5.1	
1,3-Dichlorobenzene	1.711	1.653	1.627	1.688	1.735	1.768	1.697	3.1	
1,4-Dichlorobenzene	1.809	1.769	1.608	1.661	1.692	1.686	1.704	4.3	
1,2-Dichlorobenzene	1.500	1.502	1.408	1.473	1.502	1.494	1.480	2.5	
1,2-Dibromo-3-Chloropropane	0.083	0.109	0.090	0.091	0.096	0.093	0.094	8.9	
1,2,4-Trichlorobenzene	0.913	0.860	0.855	0.923	1.001	1.014	0.928	7.3	
1,2,3-Trichlorobenzene	0.766	0.782	0.751	0.825	0.873	0.891	0.815	7.1	
1,2-Dichloroethane-d4	0.554	0.532	0.451	0.526	0.504	0.523	0.515	6.8	
Dibromofluoromethane	0.335	0.357	0.285	0.338	0.329	0.354	0.333	7.8	
Toluene-d8	1.162	1.249	1.087	1.330	1.303	1.393	1.254	9	
4-Bromofluorobenzene	0.388	0.392	0.353	0.435	0.437	0.464	0.411	9.9	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.