

Instrument ID: MSVOA_D

Daily Analysis Runlog For Sequence/QC Batch ID # VD111524

Review By	Mahesh Dadoda	Review On	11/18/2024 3:13:25 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/18/2024 3:17:20 PM
SubDirectory	VD111524	HP Acquire Method	HP Processing Method 82d111424s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP131514		
Initial Calibration Stds	VP131516,VP131518,VP131520,VP131522,VP131524,VP131526		
CCC	VP131535,VP131536		
Internal Standard/PEM	VP128297		
ICV/I.BLK	VP131528		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VD080010.D	14 Nov 2024 17:14	Method failed for Comp. #05	RP/MD	Not Ok
2	VSTDICC005	VSTDICC005	VD080011.D	14 Nov 2024 18:02	%D failed for Comp. #16 in 05PPB	RP/MD	Not Ok
3	VSTDICC010	VSTDICC010	VD080012.D	14 Nov 2024 18:29		RP/MD	Not Ok
4	VSTDICC020	VSTDICC020	VD080013.D	14 Nov 2024 18:56		RP/MD	Not Ok
5	VSTDICCC050	VSTDICCC050	VD080014.D	14 Nov 2024 19:22	Comp. #10,16 are on Quadratic Regression	RP/MD	Not Ok
6	VSTDICC100	VSTDICC100	VD080015.D	14 Nov 2024 19:49		RP/MD	Not Ok
7	VSTDICC150	VSTDICC150	VD080016.D	14 Nov 2024 20:16		RP/MD	Not Ok
8	VSTDICV050	ICVVD111524	VD080017.D	14 Nov 2024 20:43	ICV Failed for comp. #13	RP/MD	Not Ok
9	VIBLK	VIBLK	VD080018.D	14 Nov 2024 21:36		RP/MD	Not Ok
10	BFB	BFB	VD080019.D	15 Nov 2024 09:14		RP/MD	Ok
11	VSTDICC005	VSTDICC005	VD080020.D	15 Nov 2024 10:04	QR @ comp #05	RP/MD	Ok,M
12	VSTDICC010	VSTDICC010	VD080021.D	15 Nov 2024 10:31	%D fail for comp #05 in 10PPB	RP/MD	Ok,M
13	VSTDICC020	VSTDICC020	VD080022.D	15 Nov 2024 10:58		RP/MD	Ok,M
14	VSTDICCC050	VSTDICCC050	VD080023.D	15 Nov 2024 11:24		RP/MD	Ok,M
15	VSTDICC100	VSTDICC100	VD080024.D	15 Nov 2024 11:51		RP/MD	Ok,M
16	VSTDICC150	VSTDICC150	VD080025.D	15 Nov 2024 12:18		RP/MD	Ok,M
17	VSTDICV050	ICVVD111524	VD080026.D	15 Nov 2024 15:44		RP/MD	Ok,M

Instrument ID: MSVOA_D

Daily Analysis Runlog For Sequence/QC Batch ID # VD111524

Review By	Maresh Dadoda	Review On	11/18/2024 3:13:25 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/18/2024 3:17:20 PM
SubDirectory	VD111524	HP Acquire Method	HP Processing Method 82d111424s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP131514		
Initial Calibration Stds	VP131516,VP131518,VP131520,VP131522,VP131524,VP131526		
CCC	VP131535,VP131536		
Internal Standard/PEM	VP128297		
ICV/I.BLK	VP131528		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	VD1115SBL01	VBLK21	VD080027.D	15 Nov 2024 16:19		RP/MD	Ok
19	VD1115SBS01	VD1115SBS01	VD080028.D	15 Nov 2024 16:46		RP/MD	Ok,M
20	P4837-03	3TRK-STONE	VD080029.D	15 Nov 2024 17:13	vial-A	RP/MD	Ok
21	P4836-03	T3-STONE	VD080030.D	15 Nov 2024 17:40	ISTD Fail vial-A	RP/MD	ReRun
22	P4835-03	345-2-STONE	VD080031.D	15 Nov 2024 18:07	vial-A	RP/MD	Ok
23	P4834-03	SLP-1-STONE	VD080032.D	15 Nov 2024 18:34	Internal Standard and Surrogate Fail VIAL- A	RP/MD	ReRun
24	P4838-03	T1-STONE	VD080033.D	15 Nov 2024 19:01	ISTD Fail :Vial- A	RP/MD	ReRun
25	P4833-03	MH-731-VOC	VD080034.D	15 Nov 2024 19:28	Internal not spike; Surrogate fail Vial- A	RP/MD	Not Ok
26	VD1115SBSD01	VD1115SBSD01	VD080035.D	15 Nov 2024 19:55		RP/MD	Ok,M
27	VSTDCCC050	VSTD05021	VD080036.D	15 Nov 2024 20:22		RP/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_D

Daily Analysis Runlog For Sequence/QC Batch ID # VD111824

Review By	Maresh Dadoda	Review On	11/19/2024 1:21:34 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/19/2024 1:22:24 PM
SubDirectory	VD111824	HP Acquire Method	MSVOA_D HP Processing Method 82d111424s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131590
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131593,VP131594 VP128297

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VD080057.D	18 Nov 2024 09:21		RP/MD	Ok
2	VSTDCCC050	VSTDCCC050	VD080058.D	18 Nov 2024 09:51		RP/MD	Ok,M
3	VD1118SBL01	VD1118SBL01	VD080059.D	18 Nov 2024 10:27		RP/MD	Ok
4	VD1118SBS01	VD1118SBS01	VD080060.D	18 Nov 2024 10:55		RP/MD	Ok,M
5	VD1118SBSD01	VD1118SBSD01	VD080061.D	18 Nov 2024 11:22		RP/MD	Ok,M
6	P4825-01RE	STOCK-PILERE	VD080062.D	18 Nov 2024 12:08	Internal Standard Fail ;Surrogate fail VIAL- B	RP/MD	Confirms
7	P4824-01RE	T-20RE	VD080063.D	18 Nov 2024 12:35	Internal Standard Fail ;Surrogate fail VIAL- B	RP/MD	Confirms
8	P4857-03RE	VNJ-226RE	VD080064.D	18 Nov 2024 13:02	ISTD Fail vial-B	RP/MD	Confirms
9	P4859-01	NB-08-111424	VD080065.D	18 Nov 2024 13:30	vial-B	RP/MD	Ok
10	P4869-01	EO-02-11152024	VD080066.D	18 Nov 2024 13:57	vial-A	RP/MD	Ok
11	P4848-01	TP-1	VD080067.D	18 Nov 2024 14:24	Internal Standard Fail ;Surrogate fail VIAL- A	RP/MD	ReRun
12	P4852-01	COMP-1	VD080068.D	18 Nov 2024 14:51	Internal Standard Fail ;Surrogate fail VIAL- A	RP/MD	ReRun
13	P4852-02	COMP-2	VD080069.D	18 Nov 2024 15:18	Internal Standard Fail ;Surrogate fail VIAL- A	RP/MD	ReRun
14	P4852-03	COMP-3	VD080070.D	18 Nov 2024 15:45	Internal Standard Fail ;Surrogate fail VIAL- A	RP/MD	ReRun
15	P4850-01	OR-03-11142024	VD080071.D	18 Nov 2024 16:12	vial-A	RP/MD	Ok
16	P4860-01	DUP-01	VD080072.D	18 Nov 2024 16:39	vial-A	RP/MD	Ok

Instrument ID: MSVOA_D

Daily Analysis Runlog For Sequence/QC Batch ID # VD111824

Review By	Mahesh Dadoda	Review On	11/19/2024 1:21:34 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	11/19/2024 1:22:24 PM		
SubDirectory	VD111824	HP Acquire Method	MSVOA_D	HP Processing Method	82d111424s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP131590			
CCC		VP131593,VP131594			
Internal Standard/PEM		VP128297			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

17	P4860-02	PH2-BOT-001	VD080073.D	18 Nov 2024 17:06	vial-A	RP/MD	Ok
18	P4860-03	PH2-BOT-002	VD080074.D	18 Nov 2024 17:33	Internal Standard Fail vial-A	RP/MD	ReRun
19	P4860-04	PH2-BOT-003	VD080075.D	18 Nov 2024 18:00	vial-A	RP/MD	Ok
20	P4860-05	PH2-BOT-004	VD080076.D	18 Nov 2024 18:28	vial-A	RP/MD	Ok
21	P4860-06	PH2-BOT-009	VD080077.D	18 Nov 2024 18:55	vial-A comp.#81 is out of range	RP/MD	Not Ok
22	P4860-07	PH2-BOT-008	VD080078.D	18 Nov 2024 19:22	vial-A	RP/MD	Ok
23	P4860-08	PH2-BOT-007	VD080079.D	18 Nov 2024 19:49	vial-A	RP/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VD080080.D	18 Nov 2024 20:16		RP/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY111924

Review By	Maresh Dadoda	Review On	11/20/2024 1:01:03 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/20/2024 1:03:18 PM
SubDirectory	VY111924	HP Acquire Method	HP Processing Method 82y111924s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131638 VP131639,VP131640,VP131641,VP131642,VP131643,VP131644 VP131645

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020337.D	19 Nov 2024 14:40		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY020338.D	19 Nov 2024 15:49		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY020339.D	19 Nov 2024 16:12		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY020340.D	19 Nov 2024 16:35		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY020341.D	19 Nov 2024 16:57	Comp.#3 is on Linear Regression	SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY020342.D	19 Nov 2024 17:20	Comp.#11 is on Quadratic Regression	SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY020343.D	19 Nov 2024 17:42		SY/MD	Ok,M
8	VIBLK	VIBLK	VY020344.D	19 Nov 2024 18:06		SY/MD	Ok
9	VSTDICV050	ICVVY111924	VY020345.D	19 Nov 2024 18:52		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112024

Review By	Maresh Dadoda	Review On	11/21/2024 3:23:29 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/21/2024 3:24:50 PM
SubDirectory	VY112024	HP Acquire Method	MSVOA_Y HP Processing Method 82y111924s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131659
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131661,VP131662 VP128297

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020346.D	20 Nov 2024 08:23		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY020347.D	20 Nov 2024 09:34		SY/MD	Ok,M
3	VY1120SBL01	VY1120SBL01	VY020348.D	20 Nov 2024 10:07		SY/MD	Ok
4	VY1120SBS01	VY1120SBS01	VY020349.D	20 Nov 2024 10:53		SY/MD	Ok,M
5	VY1120SBSD01	VY1120SBSD01	VY020350.D	20 Nov 2024 11:16		SY/MD	Ok,M
6	P4910-07	MH-759-VOC	VY020351.D	20 Nov 2024 11:57	Internal Standard Fail vial-A	SY/MD	ReRun
7	P4910-03	MH-COTTAGE-VOC	VY020352.D	20 Nov 2024 12:20	Internal Standard Fail vial-A	SY/MD	ReRun
8	P4929-01	ARS520	VY020353.D	20 Nov 2024 12:44	Internal Standard Fail vial-A	SY/MD	ReRun
9	P4924-03	MH-4-VOC	VY020354.D	20 Nov 2024 13:07	vial-A	SY/MD	Ok
10	P4916-03	TP-1-VOC	VY020355.D	20 Nov 2024 13:30	vial-A	SY/MD	Ok
11	P4916-07	TP-2-VOC	VY020356.D	20 Nov 2024 13:54	vial-A	SY/MD	Ok
12	P4916-11	TP-3-VOC	VY020357.D	20 Nov 2024 14:17	vial-A	SY/MD	Ok
13	P4909-02	BU-02-111824	VY020358.D	20 Nov 2024 14:41	Internal Standard Fail vial-A	SY/MD	ReRun
14	P4908-01	SP-1	VY020359.D	20 Nov 2024 15:04	vial-A	SY/MD	Ok
15	P4908-03	SP-2	VY020360.D	20 Nov 2024 15:27	Internal Standard Fail vial-A	SY/MD	ReRun
16	P4870-03	TP-1-VOC	VY020361.D	20 Nov 2024 15:51	Internal Standard Fail vial-A	SY/MD	ReRun
17	P4870-12	TP-15-VOC	VY020362.D	20 Nov 2024 16:14	Internal Standard Fail vial-A	SY/MD	ReRun
18	P4870-09	MH-736-VOC	VY020363.D	20 Nov 2024 16:38	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112024

Review By	Mahesh Dadoda	Review On	11/21/2024 3:23:29 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	11/21/2024 3:24:50 PM		
SubDirectory	VY112024	HP Acquire Method	MSVOA_Y	HP Processing Method	82y111924s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP131659			
CCC		VP131661,VP131662			
Internal Standard/PEM		VP128297			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4870-06	MH-735-VOC	VY020364.D	20 Nov 2024 17:01	Internal Standard Fail vial-A	SY/MD	ReRun
20	P4849-01	RR-1	VY020365.D	20 Nov 2024 17:24	vial-A	SY/MD	Ok
21	P4887-04	MH-739-VOC	VY020366.D	20 Nov 2024 17:48	vial-A	SY/MD	Ok
22	P4887-08	MH-760-VOC	VY020367.D	20 Nov 2024 18:11	vial-A	SY/MD	Ok
23	P4860-03	PH2-BOT-002	VY020368.D	20 Nov 2024 18:35	vial-B	SY/MD	Ok
24	P4860-06	PH2-BOT-009	VY020369.D	20 Nov 2024 18:58	vial-B	SY/MD	Ok
25	P4768-04	B-4	VY020370.D	20 Nov 2024 19:22	vial-B	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY020371.D	20 Nov 2024 19:44		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112124

Review By	Maresh Dadoda	Review On	11/22/2024 3:31:25 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/22/2024 3:36:07 PM
SubDirectory	VY112124	HP Acquire Method	HP Processing Method 82y111924s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131692
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131693,VP131694 VP128297

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020372.D	21 Nov 2024 08:32		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY020373.D	21 Nov 2024 09:25		SY/MD	Ok,M
3	VY1121SBL01	VY1121SBL01	VY020374.D	21 Nov 2024 09:59		SY/MD	Ok
4	VY1121SBS01	VY1121SBS01	VY020375.D	21 Nov 2024 10:39		SY/MD	Ok,M
5	VY1121SBSD01	VY1121SBSD01	VY020376.D	21 Nov 2024 11:02		SY/MD	Ok,M
6	P4852-01	COMP-1	VY020377.D	21 Nov 2024 11:25	vial-B	SY/MD	Ok
7	P4852-02RE	COMP-2RE	VY020378.D	21 Nov 2024 11:49	ISTD Fail vial-B	SY/MD	Confirms
8	P4852-03	COMP-3	VY020379.D	21 Nov 2024 12:12	vial-B	SY/MD	Ok
9	P4908-03	SP-2	VY020380.D	21 Nov 2024 12:36	Vial-B	SY/MD	Ok
10	P4909-02	BU-02-111824	VY020381.D	21 Nov 2024 12:59	Vial-B	SY/MD	Ok
11	P4929-01RE	ARS520RE	VY020382.D	21 Nov 2024 13:23	Internal Standard Fail vial-B	SY/MD	Confirms
12	P4910-07	MH-759-VOC	VY020383.D	21 Nov 2024 13:46	vial-B	SY/MD	Ok
13	P4910-03	MH-COTTAGE-VOC	VY020384.D	21 Nov 2024 14:09	vial-B	SY/MD	Ok
14	P4870-03	TP-1-VOC	VY020385.D	21 Nov 2024 14:33	vial-B	SY/MD	Ok
15	P4870-06	MH-735-VOC	VY020386.D	21 Nov 2024 14:56	vial-B	SY/MD	Ok
16	P4870-12	TP-15-VOC	VY020387.D	21 Nov 2024 15:20	vial-B	SY/MD	Ok
17	P4892-02	WB-310-BOT	VY020388.D	21 Nov 2024 15:43	Internal Standard Fail vial-B	SY/MD	ReRun
18	P4860-10	PH2-BOT-005	VY020389.D	21 Nov 2024 16:06	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY112124

Review By	Mahesh Dadoda	Review On	11/22/2024 3:31:25 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/22/2024 3:36:07 PM
SubDirectory	VY112124	HP Acquire Method	HP Processing Method 82y111924s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP131692		
CCC	VP131693,VP131694		
Internal Standard/PEM	VP128297		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P4860-09	PH2-BOT-006	VY020390.D	21 Nov 2024 16:30	vial-A	SY/MD	Ok
20	P4925-03	MH-741-VOC	VY020391.D	21 Nov 2024 16:53	vial-A	SY/MD	Ok
21	P4893-03	MH-763-VOC	VY020392.D	21 Nov 2024 17:17	vial-A	SY/MD	Ok
22	P4893-07	MH-762-VOC	VY020393.D	21 Nov 2024 17:40	vial-A	SY/MD	Ok
23	P4925-07	MH-758-VOC	VY020394.D	21 Nov 2024 18:04	vial-A	SY/MD	Ok
24	P4892-01	WB-310-TOP	VY020395.D	21 Nov 2024 18:27	vial-A	SY/MD	Ok
25	P4938-07	MH-734-VOC	VY020396.D	21 Nov 2024 18:50	vial-A	SY/MD	Ok
26	P4938-03	MH-732-VOC	VY020397.D	21 Nov 2024 19:14	vial-A	SY/MD	Ok
27	P4936-01	PL-01-11202024	VY020398.D	21 Nov 2024 19:37	Internal Standard Fail vial-A	SY/MD	ReRun
28	VSTDCCC050	VSTDCCC050EC	VY020399.D	21 Nov 2024 20:00		SY/MD	Ok,M

M : Manual Integration