

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CORN02
 Lab Code: CHEM Case No.: P4860 SAS No.: P4860 SDG NO.: P4860
 Lab File ID: VY020347.D Date Analyzed: 11/20/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:34
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	139421	13.352				
UPPER LIMIT	278842	13.852				
LOWER LIMIT	69710.5	12.852				
EPA SAMPLE NO.						
PH2-BOT-002	102362	13.35				
PH2-BOT-009	99327	13.35				
VY1120SBL01	104132	13.35				
VY1120SBS01	135305	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.