

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CORN02
 Lab Code: CHEM Case No.: P4860 SAS No.: P4860 SDG NO.: P4860
 Lab File ID: VY020373.D Date Analyzed: 11/21/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:25
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	121826	13.352				
UPPER LIMIT	243652	13.852				
LOWER LIMIT	60913	12.852				
EPA SAMPLE NO.						
PH2-BOT-006	106466	13.35				
PH2-BOT-005	101642	13.35				
VY1121SBL01	89046	13.35				
VY1121SBS01	121310	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.