



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

AECOM

Project Name: Meeker Ave Plumes Superfund Site RI FS

Project # N/A

Chemtech Project # P4861

Test Name: TCLP Mercury, TCLP ICP Metals

A. Number of Samples and Date of Receipt:

1 Water sample was received on 11/14/2024.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Cyanide, Flash Point, Paint Filter, PCB, pH, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP Pesticide, TCLP VOA and TCLP ZHE Extraction. This data package contains results for TCLP Mercury, TCLP ICP Metals.

C. Analytical Techniques:

The analysis of TCLP ICP Metals was based on method 6010D, digestion based on method 3010 (waters). The analysis and digestion of TCLP Mercury was based on method 7470A and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____