

Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB133477

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/15/2024	17:15
2	CAL2	1	Water	NA	NA	20.2	7.00	11/15/2024	17:16
3	CAL3	1	Water	NA	NA	20.4	10.02	11/15/2024	17:20
4	ICV	1	Water	NA	NA	20.3	7.00	11/15/2024	17:25
5	CCV1	1	Water	NA	NA	20.2	2.01	11/15/2024	17:26
6	P4861-01	1	Water	NA	NA	20.1	7.37	11/15/2024	17:29
7	P4861-01DUP	1	Water	NA	NA	20.2	7.38	11/15/2024	17:30
8	P4864-01	1	Water	NA	NA	21.1	2.76	11/15/2024	17:35
9	P4867-02	1	Water	NA	NA	20.2	7.22	11/15/2024	17:40
10	P4867-03	1	Water	NA	NA	20.2	6.82	11/15/2024	17:45
11	P4871-01	1	Water	NA	NA	20.6	7.37	11/15/2024	17:48
12	P4890-01	1	Water	NA	NA	20.1	6.42	11/15/2024	17:50
13	P4890-02	1	Water	NA	NA	20.7	7.01	11/15/2024	17:52
14	P4890-03	1	Water	NA	NA	20.7	7.01	11/15/2024	17:55
15	P4890-04	1	Water	NA	NA	20.8	6.20	11/15/2024	17:59
16	CCV2	1	Water	NA	NA	20.3	12.02	11/16/2024	18:00

WORKLIST(Hardcopy Internal Chain)

VB 133477

WorkList Name : ph p4862 water WorkList ID : 185496 Department : Wet-Chemistry Date : 11-15-2024 17:05:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4861-01	WC-11-A-202411	Water	pH	Cool 4 deg C	AECO02	L41	11/13/2024	9040C
P4864-01	MONTCLAIR-TOTE-2	Water	pH	Cool 4 deg C	PSEG03	M11	11/15/2024	9040C
P4867-02	TT-070-IDWGW-20241113	Water	pH	Cool 4 deg C	TETR06	L41	11/13/2024	9040C
P4867-03	TT-071-IDWGW-20241113	Water	pH	Cool 4 deg C	TETR06	L41	11/13/2024	9040C
P4867-04	TT-072-IDWGW-20241113	Water	pH	Cool 4 deg C	TETR06	L41	11/13/2024	9040C
P4871-01	WC-10-A-202411	Water	pH	Cool 4 deg C	AECO02	M11	11/13/2024	9040C
P4890-01	#3711	Water	pH	Cool 4 deg C	PSEG03	M11	11/15/2024	9040C
P4890-02	D3617	Water	pH	Cool 4 deg C	PSEG03	M11	11/15/2024	9040C
P4890-03	D3690	Water	pH	Cool 4 deg C	PSEG03	M11	11/15/2024	9040C
P4890-04	D3721	Water	pH	Cool 4 deg C	PSEG03	M11	11/15/2024	9040C

78

78

Date/Time 11-15-24 17:10
 Raw Sample Received by: 78 wgs
 Raw Sample Relinquished by: 78 wgs

Date/Time 11-15-24 17:15-24 19:13
 Raw Sample Received by: 78 wgs
 Raw Sample Relinquished by: 78 wgs