



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: P4861

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB133474

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Flash Point	ICV ° F	82.1	81	101	78-84	11/15/2024

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: P4861

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB133475

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.095	0.099	96	85-115	11/15/2024
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	11/15/2024
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	11/15/2024
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	11/15/2024
Sample ID: CCV4 Reactive Cyanide	mg/L	0.26	0.25	104	90-110	11/15/2024

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: P4861

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB133477

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	11/15/2024
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	11/15/2024
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	11/15/2024

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: P4861

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB133498

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.095	0.099	96	90-110	11/18/2024
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	11/18/2024
Sample ID: Cyanide	CCV2	mg/L	0.24	0.25	96	90-110	11/18/2024
Sample ID: Cyanide	CCV3	mg/L	0.26	0.25	104	90-110	11/18/2024