



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: AECOM **SDG No.:** P4861
Contract: AECO02 **Lab Code:** CHEM **Case No.:** P4861 **SAS No.:** P4861
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	19.8	20.0	99	40 - 160	P	11/18/2024	13:20	LB133502
	Barium	98.9	100	99	40 - 160	P	11/18/2024	13:20	LB133502
	Cadmium	7.02	6.0	117	40 - 160	P	11/18/2024	13:20	LB133502
	Chromium	10.5	10.0	105	40 - 160	P	11/18/2024	13:20	LB133502
	Lead	11.7	12.0	97	40 - 160	P	11/18/2024	13:20	LB133502
	Selenium	21.8	20.0	109	40 - 160	P	11/18/2024	13:20	LB133502
	Silver	11.0	10.0	110	40 - 160	P	11/18/2024	13:20	LB133502
CRA	Mercury	0.17	0.2	87	40 - 160	CV	11/19/2024	11:30	LB133509
CRI01	Arsenic	19.4	20.0	97	40 - 160	P	11/19/2024	11:43	LB133524
	Barium	105	100	105	40 - 160	P	11/19/2024	11:43	LB133524
	Cadmium	5.78	6.0	96	40 - 160	P	11/19/2024	11:43	LB133524
	Chromium	10.2	10.0	102	40 - 160	P	11/19/2024	11:43	LB133524
	Lead	12.1	12.0	101	40 - 160	P	11/19/2024	11:43	LB133524
	Selenium	18.1	20.0	90	40 - 160	P	11/19/2024	11:43	LB133524
	Silver	10.1	10.0	101	40 - 160	P	11/19/2024	11:43	LB133524