

Surrogate Summary

SW-846

SDG No.: P4861

Client: AECOM

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4848-02MS	TP-1MS	2-Fluorophenol	150	140	93		10	139
		Phenol-d6	150	136	91		10	134
		Nitrobenzene-d5	100	101	101		49	133
		2-Fluorobiphenyl	100	99.0	99		52	132
		2,4,6-Tribromophenol	150	159	106		44	137
		Terphenyl-d14	100	110	110		48	125
P4848-02MSD	TP-1MSD	2-Fluorophenol	150	144	96		10	139
		Phenol-d6	150	138	92		10	134
		Nitrobenzene-d5	100	104	104		49	133
		2-Fluorobiphenyl	100	100	100		52	132
		2,4,6-Tribromophenol	150	159	106		44	137
		Terphenyl-d14	100	113	113		48	125
P4861-01	WC-11-A-202411	2-Fluorophenol	150	132	88		10	139
		Phenol-d6	150	119	79		10	134
		Nitrobenzene-d5	100	100	100		49	133
		2-Fluorobiphenyl	100	99.9	100		52	132
		2,4,6-Tribromophenol	150	158	105		44	137
		Terphenyl-d14	100	110	110		48	125
PB165019TB	PB165019TB	2-Fluorophenol	150	126	84		10	139
		Phenol-d6	150	121	81		10	134
		Nitrobenzene-d5	100	86.6	87		49	133
		2-Fluorobiphenyl	100	87.6	88		52	132
		2,4,6-Tribromophenol	150	124	83		44	137
		Terphenyl-d14	100	85.9	86		48	125
PB165052BL	PB165052BL	2-Fluorophenol	150	138	92		10	139
		Phenol-d6	150	132	88		10	134
		Nitrobenzene-d5	100	92.2	92		49	133
		2-Fluorobiphenyl	100	92.9	93		52	132
		2,4,6-Tribromophenol	150	133	88		44	137
		Terphenyl-d14	100	102	102		48	125
PB165052BS	PB165052BS	2-Fluorophenol	150	136	91		10	139
		Phenol-d6	150	133	89		10	134
		Nitrobenzene-d5	100	90.2	90		49	133
		2-Fluorobiphenyl	100	91.1	91		52	132
		2,4,6-Tribromophenol	150	143	95		44	137
		Terphenyl-d14	100	89.0	89		48	125