

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4861

Client: AECOM

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P4848-02MSD	Client Sample ID:	TP-1MSD					DataFile:	BF140466.D		
Pyridine	500	0	320	ug/L	64	0			10	109	20
1,4-Dichlorobenzene	500	0	410	ug/L	82	2			55	125	20
2-Methylphenol	500	0	510	ug/L	102	2			37	126	20
3+4-Methylphenols	500	0	520	ug/L	104	2			31	127	20
Hexachloroethane	500	0	410	ug/L	82	2			49	110	20
Nitrobenzene	500	0	460	ug/L	92	0			62	112	20
Hexachlorobutadiene	500	0	440	ug/L	88	2			52	125	20
2,4,6-Trichlorophenol	500	0	550	ug/L	110	4			78	112	20
2,4,5-Trichlorophenol	500	0	520	ug/L	104	0			71	111	20
2,4-Dinitrotoluene	500	0	540	ug/L	108	2			50	142	20
Hexachlorobenzene	500	0	540	ug/L	108	2			72	115	20
Pentachlorophenol	1000	0	1100	ug/L	110	0			25	139	20